

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020215\
 Data File : BE089463.D
 Acq On : 3 Feb 2015 1:41
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE020215

Quant Time: Feb 03 11:50:42 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020215.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 03 11:20:27 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	78	0.00
2	n-Nitrosodimethylamine	1.000	0.964	3.6	77	0.00
3 S	2-Fluorophenol	1.000	0.984	1.6	80	0.00
4 S	Phenol-d6	1.000	0.957	4.3	75	0.00
5 C	Phenol	1.000	0.890	11.0	69	0.02
6	bis(2-Chloroethyl)ether	1.000	0.950	5.0	75	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	72	0.00
8 S	Nitrobenzene-d5	1.000	0.969	3.1	71	0.00
9	Nitrobenzene	1.000	0.995	0.5	71	0.00
10	2,4-Dimethylphenol	1.000	1.002	-0.2	72	0.00
11 C	2,4-Dichlorophenol	1.000	1.002	-0.2	72	0.00
12	Hexachlorobutadiene	1.000	0.963	3.7	72	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	72	0.00
14 S	2,4,6-Tribromophenol	1.000	1.231	-23.1	85	0.00
15 S	2-Fluorobiphenyl	1.000	0.917	8.3	69	0.00
16 P	2,4-Dinitrophenol	1.000	1.024	-2.4	64	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	84	0.00
18	Hexachlorobenzene	1.000	0.912	8.8	81	0.00
19 C	Pentachlorophenol	1.000	0.976	2.4	82	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	92	0.00
21	Benzidine	1.000	1.051	-5.1	108	0.00
22 S	Terphenyl-d14	1.000	1.070	-7.0	90	0.00
23	Benzo(a)anthracene	1.000	1.033	-3.3	92	0.00
24	3,3'-Dichlorobenzidine	1.000	0.929	7.1	94	0.00
25	Chrysene	1.000	0.967	3.3	91	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.804	19.6	92	0.00
27 I	Perylene-d12	5.000	5.000	0.0	95	0.00
28	Benzo(b)fluoranthene	1.000	0.967	3.3	98	0.00
29	Benzo(k)fluoranthene	1.000	1.069	-6.9	95	0.00
30 C	Benzo(a)pyrene	1.000	0.977	2.3	99	0.00
31	Dibenzo(a,h)anthracene	1.000	0.892	10.8	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0