

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020216\
 Data File : BE091383.D
 Acq On : 2 Feb 2016 18:34
 Operator : UM/IZ
 Sample : SSTDICC010
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDICC010

Quant Time: Feb 03 00:22:07 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 02 17:23:30 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.86	152	46256	5.00	ng	-0.02
4) Naphthalene-d8	9.53	136	224125	5.00	ng	-0.02
8) Acenaphthene-d10	13.34	164	135949	5.00	ng	0.00
14) Phenanthrene-d10	16.08	188	323650	5.00	ng	0.00
18) Chrysene-d12	20.22	240	396030	5.00	ng	0.00
24) Perylene-d12	22.09	264	367002	5.00	ng	0.00

System Monitoring Compounds						
2) 2-Fluorophenol	4.90	112	111778	22.68	ng	-0.04
3) Phenol-d6	6.51	99	159107	20.20	ng	-0.05
5) Nitrobenzene-d5	8.12	82	155785	11.99	ng	-0.04
9) 2,4,6-Tribromophenol	14.89	330	68087	24.41	ng	0.00
10) 2-Fluorobiphenyl	11.98	172	371633	12.65	ng	-0.03
20) Terphenyl-d14	18.61	244	629708	16.76	ng	-0.02

Target Compounds						Qvalue
6) Naphthalene	9.57	128	482977	10.10	ng	99
7) 2-Methylnaphthalene	11.10	142	318820	17.05	ng	94
11) Acenaphthylene	13.09	152	2426979	13.62	ng	99
12) Acenaphthene	13.41	154	358522	9.29	ng	# 68
13) Fluorene	14.39	166	492216	11.49	ng	95
15) Phenanthrene	16.12	178	802706	10.02	ng	99
16) Anthracene	16.20	178	815003	14.52	ng	99
17) Fluoranthene	18.08	202	898212	15.54	ng	99
19) Pyrene	18.46	202	959515	15.87	ng	100
21) Benzo(a)anthracene	20.18	228	898952	16.31	ng	99
22) Chrysene	20.24	228	961672	10.00	ng	97
23) Indeno(1,2,3-cd)pyrene	23.69	276	917007	14.60	ng	93
25) Benzo(b)fluoranthene	21.54	252	927813	15.06	ng	96
26) Benzo(k)fluoranthene	21.57	252	908831	12.62	ng	# 96
27) Benzo(a)pyrene	22.00	252	861994	13.70	ng	# 97
28) Dibenzo(a,h)anthracene	23.64	278	867869	18.57	ng	# 96
29) Benzo(g,h,i)perylene	24.24	276	928610	15.60	ng	# 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020216\
Data File : BE091383.D
Acq On : 2 Feb 2016 18:34
Operator : UM/IZ
Sample : SSTDIC010
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
SSTDIC010

Quant Time: Feb 03 00:22:07 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 02 17:23:30 2016
Response via : Initial Calibration

