

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020218\
 Data File : BE095411.D
 Acq On : 2 Feb 2018 23:34
 Operator : SJ/JU
 Sample : J1317-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Feb 03 05:27:42 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 03 05:26:35 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	1860	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	5544	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	3671	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	7595	0.40	ng/ul	0.00
16) Chrysene-d12	21.81	240	7445	0.40	ng/ul	0.00
20) Perylene-d12	24.27	264	421493	0.40	ng/ul	-0.16
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	3729	0.41	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	10225	0.35	ng/ul	0.00
Target Compounds				Ovalue		
3) Naphthalene	11.31	128	460	0.030	ng/ul#	62
8) Acenaphthene	15.07	153	1802	0.132	ng/ul#	25
17) Pyrene	20.05	202	1433	0.057	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020218\
Data File : BE095411.D
Acq On : 2 Feb 2018 23:34
Operator : SJ/JU
Sample : J1317-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Feb 03 05:27:42 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 03 05:26:35 2018
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.030 ng/ul

RT: 11.31 min Scan# 1348

Delta R.T. 0.00 min

Lab File: BE095411.D

Acq: 2 Feb 2018 23:34

Instrument :

BNA_E

ClientSampleId :


Tot Ion:128 Resp: 460

Ion Ratio Lower Upper

128 100

129 25.9 11.3 16.9#




127 26.5 4.0 6.0#



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 129.00 (128.70 to 129.70): E

Ion 127.00 (126.70 to 127.70): E



#8

Acenaphthene

Concen: 0.132 ng/ul

RT: 15.07 min Scan# 1824

Delta R.T. -0.01 min

Lab File: BE095411.D

Acq: 2 Feb 2018 23:34


Tgt Ion:153 Resp: 1802

Ion Ratio Lower Upper

153 100

152 5.1 36.0 54.0#



154 12.7 72.0 108.0#




Abundance Ion 153.00 (152.70 to 153.70): E

Ion 152.00 (151.70 to 152.70): E

Ion 154.00 (153.70 to 154.70): E





#17
 Pvrene
 Concen: 0.057 ng/ul
 RT: 20.05 min Scan# 2460
 Delta R.T. -0.03 min
 Lab File: BE095411.D
 Acq: 2 Feb 2018 23:34

Instrument :
 BNA_E
 ClientSampleId :

Tot Ion:202 Resp: 1433
 Ion Ratio Lower Upper
 202 100
 101 160.8 18.2 27.4#
 100 1022.4 15.6 23.4#

