

Data Path : U:\HPCHEM1\BNA E\DATA\BE020218\
 Data File : BE095414.D
 Acq On : 3 Feb 2018 1:19
 Operator : SJ/JU
 Sample : J1296-18
 Misc :
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C75

Manual Integrations
 APPROVED

Sohil
 2/5/2018 5:48:35 PM

Quant Time: Feb 03 05:46:31 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 03 05:26:35 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.44	152	1547m	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	6580	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	4300	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	8745	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	8506	0.40	ng/ul	0.00
20) Perylene-d12	24.44	264	7729m	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	3499	0.33	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	8756	0.26	ng/ul	0.00
Target Compounds						
11) Pentachlorophenol	17.36	266	93	0.055	ng/ul	Ovalue 94
12) Phenanthrene	17.75	178	940m	0.032	ng/ul	

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : U:\HPCHEM1\BNA E\DATA\BE020218\
Data File : BE095414.D
Acq On : 3 Feb 2018 1:19
Operator : SJ/JU
Sample : J1296-18
Misc :
ALS Vial : 22 Sample Multiplier: 1


Instrument :
BNA_E
ClientSampleId :
F6C75

Manual Integrations
APPROVED

Sohil
2/5/2018 5:48:35 PM

Quant Time: Feb 03 05:46:31 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 03 05:26:35 2018
Response via : Initial Calibration









#11
 Pentachlorophenol
 Concen: 0.055 ng/ul
 RT: 17.36 min Scan# 1997
 Delta R.T. -0.00 min
 Lab File: BE095414.D
 Acq: 3 Feb 2018 1:19

Instrument :
 BNA_E
 ClientSampled :
 F6C75

Tot Ion	Ratio	Lower	Upper
266	100		
264	54.8	47.9	71.9
268	66.7	49.8	74.8

Manual Integrations
 APPROVED

Sohil
 2/5/2018 5:48:35 PM



#12
 Phenanthrene
 Concen: 0.032 ng/ul m
 RT: 17.75 min Scan# 2014
 Delta R.T. -0.00 min
 Lab File: BE095414.D
 Acq: 3 Feb 2018 1:19

Tgt Ion	Ratio	Lower	Upper
178	100		
179	24.7	18.4	27.6
176	25.3	13.6	20.4

