

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089466.D
 Acq On : 3 Feb 2015 16:19
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC005

Quant Time: Feb 04 00:30:36 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:22:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.53	152	58239	5.00	ng	0.00
7) Naphthalene-d8	12.43	136	231656	5.00	ng	0.00
13) Acenaphthene-d10	16.61	164	99385	5.00	ng	0.00
17) Phenanthrene-d10	19.93	188	188736	5.00	ng	0.00
20) Chrysene-d12	23.79	240	188517	5.00	ng	0.00
27) Perylene-d12	25.49	264	177028	5.00	ng	0.00

System Monitoring Compounds						
3) 2-Fluorophenol	6.72	112	94321	4.84	ng	0.00
4) Phenol-d6	8.84	99	117879	5.15	ng	0.00
8) Nitrobenzene-d5	10.79	82	91448	5.41	ng	0.00
14) 2,4,6-Tribromophenol	18.52	330	20434	7.87	ng	0.00
15) 2-Fluorobiphenyl	15.06	172	130008	4.69	ng	0.00
22) Terphenyl-d14	22.43	244	139043	6.40	ng	0.00


Target Compounds				Qvalue		
2) n-Nitrosodimethylamine	3.54	42	58283	6.23	ng	# 76
5) Phenol	8.87	94	116909	4.76	ng	# 67
6) bis(2-Chloroethyl)ether	9.01	93	91972	5.13	ng	# 100
9) Nitrobenzene	10.83	77	104247	5.43	ng	99
10) 2,4-Dimethylphenol	11.78	122	74261	5.40	ng	95
11) 2,4-Dichlorophenol	12.16	162	56439	5.44	ng	86
12) Hexachlorobutadiene	12.81	225	34641	4.74	ng	99
16) 2,4-Dinitrophenol	16.85	184	5567	9.31	ng	# 62
18) Hexachlorobenzene	19.17	284	47648	4.81	ng	# 86
19) Pentachlorophenol	19.60	266	13448	11.55	ng	99
21) Benzidine	22.09	184	44752	3.33	ng	97
23) Benzo(a)anthracene	23.77	228	878127	5.77	ng	99
24) 3,3'-Dichlorobenzidine	23.76	252	76812	5.74	ng	97
25) Chrysene	23.82	228	844663	4.63	ng	97
26) Indeno(1,2,3-cd)pyrene	26.97	276	243053	4.55	ng	99
28) Benzo(b)fluoranthene	25.04	252	204174	8.45	ng	99
29) Benzo(k)fluoranthene	25.08	252	229176	5.16	ng	99
30) Benzo(a)pyrene	25.42	252	190845	7.04	ng	98
31) Dibenzo(a,h)anthracene	27.00	278	208140	6.41	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
Data File : BE089466.D
Acq On : 3 Feb 2015 16:19
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC005

Quant Time: Feb 04 00:30:36 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 04 00:22:17 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 9.53 min Scan# 1125

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion:152 Resp: 58239

Ion Ratio Lower Upper

152 100

150 153.1 122.1 183.1




115 68.8 53.2 79.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

n-Nitrosodimethylamine

Concen: 6.23 ng

RT: 3.54 min Scan# 316

Delta R.T. -0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion: 42 Resp: 58283

Ion Ratio Lower Upper

42 100

74 103.8 65.4 98.2#



44 6.0 11.7 17.5#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#3

2-Fluorophenol

Concen: 4.84 ng

RT: 6.72 min Scan# 785

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDICC005


Tgt Ion: 112 Resp: 94321

Ion Ratio Lower Upper

112 100

64 64.7 47.0 70.6




63 34.4 24.2 36.2



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#4

Phenol-d6

Concen: 5.15 ng

RT: 8.84 min Scan# 1047

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion: 99 Resp: 117879

Ion Ratio Lower Upper

99 100

42 22.2 11.8 17.6#



71 40.7 28.7 43.1




Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#5

Phenol

Concen: 4.76 ng

RT: 8.87 min Scan# 1051

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion: 94 Resp: 116909

Ion Ratio Lower Upper

94 100

65 90.7 33.9 73.9#




66 34.6 19.1 59.1



Abundance Ion 94.00 (93.70 to 94.70): BE0

Ion 65.00 (64.70 to 65.70): BE0

Ion 66.00 (65.70 to 66.70): BE0



#6

bis(2-Chloroethyl)ether

Concen: 5.13 ng

RT: 9.01 min Scan# 1066

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion: 93 Resp: 91972

Ion Ratio Lower Upper

93 100

63 70.9 57.0 85.6



95 0.0 0.0 0.0




Abundance Ion 93.00 (92.70 to 93.70): BE0

Ion 63.00 (62.70 to 63.70): BE0

Ion 95.00 (94.70 to 95.70): BE0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 12.43 min Scan# 1556

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion: 136 Resp: 231656

Ion Ratio Lower Upper

136 100

137 10.8 8.6 13.0




54 7.8 5.4 8.0



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene-d5

Concen: 5.41 ng

RT: 10.79 min Scan# 1337

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion: 82 Resp: 91448

Ion Ratio Lower Upper

82 100

128 39.8 30.3 45.5



54 49.2 43.1 64.7




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 5.43 ng

RT: 10.83 min Scan# 1346

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion: 77 Resp: 104247

Ion Ratio Lower Upper




77 100

123 42.2 33.4 50.0

65 14.1 10.9 16.3



Abundance Ion 77.00 (76.70 to 77.70): BE0
Ion 123.00 (122.70 to 123.70): BE0
Ion 65.00 (64.70 to 65.70): BE0



#10

2,4-Dimethylphenol

Concen: 5.40 ng

RT: 11.78 min Scan# 1489

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion: 122 Resp: 74261

Ion Ratio Lower Upper



122 100


107 114.8 97.3 145.9

121 55.7 46.4 69.6



Abundance Ion 122.00 (121.70 to 122.70): E
Ion 107.00 (106.70 to 107.70): E
Ion 121.00 (120.70 to 121.70): E





#11

2,4-Dichlorophenol

Concen: 5.44 ng

RT: 12.16 min Scan# 1528

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion:162 Resp: 56439

Ion Ratio Lower Upper

162 100

164 64.8 39.9 79.9

98 35.4 31.3 71.3






Abundance Ion 162.00 (161.70 to 162.70): E

Ion 164.00 (163.70 to 164.70): E

Ion 98.00 (97.70 to 98.70): E

Time-->



#12

Hexachlorobutadiene

Concen: 4.74 ng

RT: 12.81 min Scan# 1595

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:225 Resp: 34641

Ion Ratio Lower Upper

225 100

223 63.3 49.4 74.0

227 63.7 51.2 76.8






Abundance Ion 225.00 (224.70 to 225.70): E

Ion 223.00 (222.70 to 223.70): E

Ion 227.00 (226.70 to 227.70): E

Time-->





#13

Acenaphthene-d10

Concen: 5.00 ng

RT: 16.61 min Scan# 2745

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDICC005


Tgt Ion:164 Resp: 99385

Ion Ratio Lower Upper




164 100

162 98.1 77.6 116.4

160 45.2 35.6 53.4



Abundance Ion 164.00 (163.70 to 164.70): E
Ion 162.00 (161.70 to 162.70): E
Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 7.87 ng

RT: 18.52 min Scan# 3026

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:330 Resp: 20434

Ion Ratio Lower Upper



330 100


332 95.8 78.0 117.0

141 42.3 33.8 50.6



Abundance Ion 329.80 (329.50 to 330.50): E
Ion 331.80 (331.50 to 332.50): E
Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 4.69 ng
RT: 15.06 min Scan# 2397
Delta R.T. 0.00 min
Lab File: BE089466.D
Acq: 3 Feb 2015 16:19




Instrument :
BNA_E
LabSampleId :
SSTDIC005


Tgt Ion:172 Resp: 130008
Ion Ratio Lower Upper
172 100
171 33.7 26.8 40.2
170 22.0 17.7 26.5



#16
2,4-Dinitrophenol
Concen: 9.31 ng
RT: 16.85 min Scan# 2780
Delta R.T. 0.01 min
Lab File: BE089466.D
Acq: 3 Feb 2015 16:19

Tgt Ion:184 Resp: 5567
Ion Ratio Lower Upper
184 100
63 69.2 92.1 138.1#
154 49.9 61.4 92.2#





#17

Phenanthrene-d10

Concen: 5.00 ng

RT: 19.93 min Scan# 3231

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion:188 Resp: 188736

Ion Ratio Lower Upper

188 100

94 14.8 11.1 16.7




80 15.8 11.8 17.6



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#18

Hexachlorobenzene

Concen: 4.81 ng

RT: 19.17 min Scan# 3119

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:284 Resp: 47648

Ion Ratio Lower Upper

284 100

142 39.5 24.4 36.6#



249 27.7 18.2 27.2#




Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E





#19

Pentachlorophenol

Concen: 11.55 ng

RT: 19.60 min Scan# 3183

Delta R.T. -0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005


Tgt Ion:266 Resp: 13448

Ion Ratio Lower Upper




266 100

268 62.1 50.0 75.0

264 64.4 50.6 75.8



Abundance Ion 265.70 (265.40 to 266.40): E
Ion 268.00 (267.70 to 268.70): E
Ion 264.00 (263.70 to 264.70): E



#20

Chrysene-d12

Concen: 5.00 ng

RT: 23.79 min Scan# 3866

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:240 Resp: 188517

Ion Ratio Lower Upper



240 100


120 6.3 13.8 20.6#

236 22.5 20.2 30.2



Abundance Ion 240.00 (239.70 to 240.70): E
Ion 120.00 (119.70 to 120.70): E
Ion 236.00 (235.70 to 236.70): E









#21
Benzidine
Concen: 3.33 ng
RT: 22.09 min Scan# 3588
Delta R.T. -0.00 min
Lab File: BE089466.D
Acq: 3 Feb 2015 16:19




Instrument :
BNA_E
LabSampleId :
SSTDIC005


Tgt Ion:184 Resp: 44752
Ion Ratio Lower Upper
184 100
185 13.8 12.0 18.0
183 11.9 9.0 13.4



#22
Terphenyl-d14
Concen: 6.40 ng
RT: 22.43 min Scan# 3662
Delta R.T. 0.00 min
Lab File: BE089466.D
Acq: 3 Feb 2015 16:19

Tgt Ion:244 Resp: 139043
Ion Ratio Lower Upper
244 100
212 6.8 5.5 8.3
122 14.5 12.6 19.0





#23

Benzo(a)anthracene

Concen: 5.77 ng

RT: 23.77 min Scan# 3864

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDICC005


Tgt Ion:228 Resp: 878127

Ion Ratio Lower Upper

228 100

226 25.2 19.6 29.4




229 19.8 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

3,3'-Dichlorobenzidine

Concen: 5.74 ng

RT: 23.76 min Scan# 3862

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:252 Resp: 76812

Ion Ratio Lower Upper

252 100

254 59.0 48.6 73.0



126 17.3 12.4 18.6




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#25

Chrysene

Concen: 4.63 ng

RT: 23.82 min Scan# 3869

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDICC005


Tgt Ion:228 Resp: 844663

Ion Ratio Lower Upper

228 100

226 28.9 21.8 32.8




229 18.9 16.2 24.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Indeno(1,2,3-cd)pyrene

Concen: 4.55 ng

RT: 26.97 min Scan# 4493

Delta R.T. 0.01 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:276 Resp: 243053

Ion Ratio Lower Upper

276 100

138 31.0 24.2 36.4



277 24.9 19.9 29.9




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#27
Perylene-d12
 Concen: 5.00 ng
 RT: 25.49 min Scan# 4171
 Delta R.T. 0.00 min
 Lab File: BE089466.D
 Acq: 3 Feb 2015 16:19




Instrument :
 BNA_E
 LabSampleId :
 SSTDICC005


Tgt Ion: 264 Resp: 177028
 Ion Ratio Lower Upper
 264 100
 260 23.7 19.1 28.7
 265 21.6 17.0 25.6



#28
Benzo(b)fluoranthene
 Concen: 8.45 ng
 RT: 25.04 min Scan# 4073
 Delta R.T. 0.00 min
 Lab File: BE089466.D
 Acq: 3 Feb 2015 16:19

Tgt Ion: 252 Resp: 204174
 Ion Ratio Lower Upper
 252 100
 253 21.5 17.4 26.2
 125 13.2 11.3 16.9





#29

Benzo(k)fluoranthene

Concen: 5.16 ng

RT: 25.08 min Scan# 4080

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDICC005


Tgt Ion:252 Resp: 229176

Ion Ratio Lower Upper




252 100

253 22.2 17.9 26.9

125 12.7 10.6 15.8



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E



#30

Benzo(a)pyrene

Concen: 7.04 ng

RT: 25.42 min Scan# 4156

Delta R.T. 0.00 min

Lab File: BE089466.D

Acq: 3 Feb 2015 16:19


Tgt Ion:252 Resp: 190845

Ion Ratio Lower Upper



252 100


253 22.0 18.2 27.2

125 14.1 12.2 18.4



Abundance Ion 252.00 (251.70 to 252.70): E
Ion 253.00 (252.70 to 253.70): E
Ion 125.00 (124.70 to 125.70): E





#31

Dibenzo(a,h)anthracene

Concen: 6.41 ng

RT: 27.00 min Scan# 4499

Delta R.T. 0.01 min

Lab File: BE089466.D


Acq: 3 Feb 2015 16:19

Instrument :

BNA_E

LabSampleId :

SSTDIC005




Tgt Ion: 278 Resp: 208140

Ion Ratio Lower Upper

278 100

139 19.4 16.8 25.2

279 23.8 19.4 29.0



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

