

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089471.D
 Acq On : 3 Feb 2015 19:36
 Operator : TP/IZ
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.2

Quant Time: Feb 04 00:32:59 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:22:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.53	152	57169	5.00	ng	0.00
7) Naphthalene-d8	12.43	136	225647	5.00	ng	0.00
13) Acenaphthene-d10	16.61	164	96710	5.00	ng	0.00
17) Phenanthrene-d10	19.93	188	186273	5.00	ng	0.00
20) Chrysene-d12	23.78	240	180368	5.00	ng	0.00
27) Perylene-d12	25.49	264	170057	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) 2-Fluorophenol	6.72	112	3829	0.20	ng	0.00
4) Phenol-d6	8.83	99	4834	0.22	ng	0.00
8) Nitrobenzene-d5	10.79	82	3377	0.21	ng	0.00
14) 2,4,6-Tribromophenol	18.51	330	569	0.23	ng	0.00
15) 2-Fluorobiphenyl	15.06	172	5245	0.19	ng	0.00
22) Terphenyl-d14	22.43	244	5031	0.24	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) n-Nitrosodimethylamine	3.60	42	1863	0.20	ng	# 76
5) Phenol	8.86	94	4840	0.20	ng	# 61
6) bis(2-Chloroethyl)ether	9.01	93	3946	0.22	ng	# 99
9) Nitrobenzene	10.83	77	3834	0.21	ng	99
10) 2,4-Dimethylphenol	11.77	122	2797	0.21	ng	97
11) 2,4-Dichlorophenol	12.15	162	2034	0.20	ng	94
12) Hexachlorobutadiene	12.81	225	1530	0.22	ng	96
16) 2,4-Dinitrophenol	16.84	184	17	0.03	ng	# 70
18) Hexachlorobenzene	19.16	284	1943	0.20	ng	# 71
19) Pentachlorophenol	19.60	266	97	0.08	ng	# 64
21) Benzidine	22.09	184	927	0.07	ng	# 83
23) Benzo(a)anthracene	23.77	228	31508	0.22	ng	98
24) 3,3'-Dichlorobenzidine	23.76	252	2092	0.16	ng	# 95
25) Chrysene	23.82	228	35190	0.20	ng	97
26) Indeno(1,2,3-cd)pyrene	26.96	276	5374	0.11	ng	# 86
28) Benzo(b)fluoranthene	25.07	252	7668	0.33	ng	98
29) Benzo(k)fluoranthene	25.07	252	7668	0.18	ng	# 96
30) Benzo(a)pyrene	25.42	252	4123	0.16	ng	# 90
31) Dibenzo(a,h)anthracene	26.99	278	3099	0.10	ng	# 88


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
Data File : BE089471.D
Acq On : 3 Feb 2015 19:36
Operator : TP/IZ
Sample : SSTDICC0.2
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.2

Quant Time: Feb 04 00:32:59 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 04 00:22:17 2015
Response via : Initial Calibration





#1

1,4-Dichlorobenzene-d4

Concen: 5.00 ng

RT: 9.53 min Scan# 1125

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2


Tgt Ion:152 Resp: 57169

Ion Ratio Lower Upper

152 100

150 149.1 122.1 183.1




115 65.2 53.2 79.8



Abundance Ion 152.00 (151.70 to 152.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 115.00 (114.70 to 115.70): E



#2

n-Nitrosodimethylamine

Concen: 0.20 ng

RT: 3.60 min Scan# 324

Delta R.T. 0.04 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion: 42 Resp: 1863

Ion Ratio Lower Upper

42 100

74 94.6 65.4 98.2



44 45.9 11.7 17.5#




Abundance Ion 42.00 (41.70 to 42.70): BE0

Ion 74.00 (73.70 to 74.70): BE0

Ion 44.00 (43.70 to 44.70): BE0





#3

2-Fluorophenol

Concen: 0.20 ng

RT: 6.72 min Scan# 784

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2


Tgt Ion: 112 Resp: 3829

Ion Ratio Lower Upper

112 100

64 63.1 47.0 70.6




63 34.5 24.2 36.2



Abundance Ion 112.00 (111.70 to 112.70): BE0

Ion 64.00 (63.70 to 64.70): BE0

Ion 63.00 (62.70 to 63.70): BE0



#4

Phenol-d6

Concen: 0.22 ng

RT: 8.83 min Scan# 1046

Delta R.T. -0.01 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion: 99 Resp: 4834

Ion Ratio Lower Upper

99 100

42 27.8 11.8 17.6#



71 40.9 28.7 43.1






Abundance Ion 99.00 (98.70 to 99.70): BE0

Ion 42.00 (41.70 to 42.70): BE0

Ion 71.00 (70.70 to 71.70): BE0





#5

Phenol

Concen: 0.20 ng

RT: 8.86 min Scan# 1050

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2





Tgt Ion: 94 Resp: 4840

Ion Ratio Lower Upper

94 100

65 100.2 33.9 73.9#

66 36.2 19.1 59.1



#6

bis(2-Chloroethyl)ether

Concen: 0.22 ng

RT: 9.01 min Scan# 1066

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion: 93 Resp: 3946


Ion Ratio Lower Upper

93 100

63 70.3 57.0 85.6

95 0.0 0.0 0.0





#7

Naphthalene-d8

Concen: 5.00 ng

RT: 12.43 min Scan# 1556

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2


Tgt Ion: 136 Resp: 225647

Ion Ratio Lower Upper

136 100

137 11.0 8.6 13.0




54 6.4 5.4 8.0



Abundance Ion 136.00 (135.70 to 136.70): E

Ion 137.00 (136.70 to 137.70): E

Ion 54.00 (53.70 to 54.70): BE0



#8

Nitrobenzene-d5

Concen: 0.21 ng

RT: 10.79 min Scan# 1336

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion: 82 Resp: 3377

Ion Ratio Lower Upper

82 100

128 38.8 30.3 45.5



54 53.0 43.1 64.7




Abundance Ion 82.00 (81.70 to 82.70): BE0

Ion 128.00 (127.70 to 128.70): E

Ion 54.00 (53.70 to 54.70): BE0





#9

Nitrobenzene

Concen: 0.21 ng

RT: 10.83 min Scan# 1346

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2


Tgt Ion: 77 Resp: 3834

Ion Ratio Lower Upper

77 100

123 41.3 33.4 50.0




65 14.9 10.9 16.3



Abundance Ion 77.00 (76.70 to 77.70): BE089471.D (-1236) (-)

Ion 123.00 (122.70 to 123.70): BE089471.D (-1236) (-)

Ion 65.00 (64.70 to 65.70): BE089471.D (-1236) (-)



#10

2,4-Dimethylphenol

Concen: 0.21 ng

RT: 11.77 min Scan# 1488

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion: 122 Resp: 2797

Ion Ratio Lower Upper

122 100

107 124.9 97.3 145.9



121 60.9 46.4 69.6




Abundance Ion 122.00 (121.70 to 122.70): BE089471.D (-1469) (-)

Ion 107.00 (106.70 to 107.70): BE089471.D (-1469) (-)

Ion 121.00 (120.70 to 121.70): BE089471.D (-1469) (-)









#11
 2,4-Dichlorophenol
 Concen: 0.20 ng
 RT: 12.15 min Scan# 1527
 Delta R.T. 0.00 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.2

Tgt Ion:162 Resp: 2034
 Ion Ratio Lower Upper
 162 100
 164 65.7 39.9 79.9
 98 54.2 31.3 71.3




Abundance Ion 162.00 (161.70 to 162.70): E
 Ion 164.00 (163.70 to 164.70): E
 Ion 98.00 (97.70 to 98.70): E






#12
 Hexachlorobutadiene
 Concen: 0.22 ng
 RT: 12.81 min Scan# 1595
 Delta R.T. 0.00 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36

Tgt Ion:225 Resp: 1530
 Ion Ratio Lower Upper
 225 100
 223 63.7 49.4 74.0
 227 60.1 51.2 76.8



Abundance Ion 225.00 (224.70 to 225.70): E
 Ion 223.00 (222.70 to 223.70): E
 Ion 227.00 (226.70 to 227.70): E





#13

Acenaphthene-d10

Concen: 5.00 ng

RT: 16.61 min Scan# 2745

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2


Tgt Ion:164 Resp: 96710

Ion Ratio Lower Upper

164 100

162 95.4 77.6 116.4




160 43.0 35.6 53.4



Abundance Ion 164.00 (163.70 to 164.70): E

Ion 162.00 (161.70 to 162.70): E

Ion 160.00 (159.70 to 160.70): E



#14

2,4,6-Tribromophenol

Concen: 0.23 ng

RT: 18.51 min Scan# 3025

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion:330 Resp: 569

Ion Ratio Lower Upper

330 100

332 95.6 78.0 117.0



141 51.5 33.8 50.6#




Abundance Ion 329.80 (329.50 to 330.50): E

Ion 331.80 (331.50 to 332.50): E

Ion 141.00 (140.70 to 141.70): E









#15
2-Fluorobiphenyl
Concen: 0.19 ng
RT: 15.06 min Scan# 2395
Delta R.T. -0.00 min
Lab File: BE089471.D
Acq: 3 Feb 2015 19:36




Instrument :
BNA_E
LabSampleId :
SSTDIC0.2


Tgt Ion:172 Resp: 5245
Ion Ratio Lower Upper
172 100
171 34.1 26.8 40.2
170 24.3 17.7 26.5



#16
2,4-Dinitrophenol
Concen: 0.03 ng
RT: 16.84 min Scan# 2779
Delta R.T. 0.00 min
Lab File: BE089471.D
Acq: 3 Feb 2015 19:36

Tgt Ion:184 Resp: 17
Ion Ratio Lower Upper
184 100
63 150.0 92.1 138.1#
154 100.0 61.4 92.2#





#17

Phenanthrene-d10

Concen: 5.00 ng

RT: 19.93 min Scan# 3231

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2


Tgt Ion:188 Resp: 186273

Ion Ratio Lower Upper

188 100

94 13.4 11.1 16.7




80 13.7 11.8 17.6



Abundance Ion 188.00 (187.70 to 188.70): E

Ion 94.00 (93.70 to 94.70): BE0

Ion 80.00 (79.70 to 80.70): BE0



#18

Hexachlorobenzene

Concen: 0.20 ng

RT: 19.16 min Scan# 3118

Delta R.T. -0.01 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion:284 Resp: 1943

Ion Ratio Lower Upper

284 100

142 48.2 24.4 36.6#



249 34.1 18.2 27.2#




Abundance Ion 284.00 (283.70 to 284.70): E

Ion 142.00 (141.70 to 142.70): E

Ion 249.00 (248.70 to 249.70): E









#19
 Pentachlorophenol
 Concen: 0.08 ng
 RT: 19.60 min Scan# 3182
 Delta R.T. -0.01 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36




Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.2


Tgt Ion: 266 Resp: 97
 Ion Ratio Lower Upper
 266 100
 268 82.7 50.0 75.0#
 264 98.1 50.6 75.8#



#20
 Chrysene-d12
 Concen: 5.00 ng
 RT: 23.78 min Scan# 3866
 Delta R.T. -0.00 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36

Tgt Ion: 240 Resp: 180368
 Ion Ratio Lower Upper
 240 100
 120 18.4 13.8 20.6
 236 25.8 20.2 30.2





#21

Benzidine

Concen: 0.07 ng

RT: 22.09 min Scan# 3589

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2





Tgt Ion:184 Resp: 927

Ion Ratio Lower Upper

184 100

185 21.2 12.0 18.0#

183 18.5 9.0 13.4#



#22

Terphenyl-d14

Concen: 0.24 ng

RT: 22.43 min Scan# 3662

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36




Tgt Ion:244 Resp: 5031


Ion Ratio Lower Upper

244 100

212 8.1 5.5 8.3

122 17.8 12.6 19.0





#23

Benzo(a)anthracene

Concen: 0.22 ng

RT: 23.77 min Scan# 3865

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2


Tgt Ion:228 Resp: 31508

Ion Ratio Lower Upper

228 100

226 25.7 19.6 29.4




229 19.3 15.9 23.9



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#24

3,3'-Dichlorobenzidine

Concen: 0.16 ng

RT: 23.76 min Scan# 3863

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion:252 Resp: 2092

Ion Ratio Lower Upper

252 100

254 59.1 48.6 73.0



126 21.6 12.4 18.6#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E





#25

Chrysene

Concen: 0.20 ng

RT: 23.82 min Scan# 3870

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2


Tgt Ion:228 Resp: 35190

Ion Ratio Lower Upper

228 100

226 28.3 21.8 32.8




229 22.0 16.2 24.2



Abundance Ion 228.00 (227.70 to 228.70): E

Ion 226.00 (225.70 to 226.70): E

Ion 229.00 (228.70 to 229.70): E



#26

Indeno(1,2,3-cd)pyrene

Concen: 0.11 ng

RT: 26.96 min Scan# 4492

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion:276 Resp: 5374

Ion Ratio Lower Upper

276 100

138 21.2 24.2 36.4#



277 19.5 19.9 29.9#




Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E









#27
Perylene-d12
 Concen: 5.00 ng
 RT: 25.49 min Scan# 4172
 Delta R.T. 0.00 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36




Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.2


Tgt Ion:264 Resp: 170057
 Ion Ratio Lower Upper
 264 100
 260 23.6 19.1 28.7
 265 21.7 17.0 25.6



#28
Benzo(b)fluoranthene
 Concen: 0.33 ng
 RT: 25.07 min Scan# 4080
 Delta R.T. 0.03 min
 Lab File: BE089471.D
 Acq: 3 Feb 2015 19:36

Tgt Ion:252 Resp: 7668
 Ion Ratio Lower Upper
 252 100
 253 21.8 17.4 26.2
 125 16.4 11.3 16.9





#29

Benzo(k)fluoranthene

Concen: 0.18 ng

RT: 25.07 min Scan# 4080

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDIC0.2


Tgt Ion:252 Resp: 7668

Ion Ratio Lower Upper

252 100

253 21.8 17.9 26.9




125 16.4 10.6 15.8#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#30

Benzo(a)pyrene

Concen: 0.16 ng

RT: 25.42 min Scan# 4156

Delta R.T. -0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36


Tgt Ion:252 Resp: 4123

Ion Ratio Lower Upper

252 100

253 24.2 18.2 27.2



125 23.9 12.2 18.4#




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E





#31

Dibenzo(a,h)anthracene

Concen: 0.10 ng

RT: 26.99 min Scan# 4499

Delta R.T. 0.00 min

Lab File: BE089471.D

Acq: 3 Feb 2015 19:36

Instrument :

BNA_E

LabSampleId :

SSTDICC0.2


Tgt Ion: 278 Resp: 3099

Ion Ratio Lower Upper

278 100

139 27.8 16.8 25.2#

279 29.4 19.4 29.0#



Abundance Ion 278.00 (277.70 to 278.70): E

Ion 139.00 (138.70 to 139.70): E

Ion 279.00 (278.70 to 279.70): E

