

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089472.D
 Acq On : 3 Feb 2015 20:15
 Operator : TP/IZ
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

apatel
 2/4/2015 2:55:13 PM

Quant Time: Feb 04 00:53:56 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:22:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.53	152	59178	5.00	ng	0.00
7) Naphthalene-d8	12.42	136	233815	5.00	ng	0.00
13) Acenaphthene-d10	16.61	164	100560	5.00	ng	0.00
17) Phenanthrene-d10	19.93	188	195073	5.00	ng	0.00
20) Chrysene-d12	23.78	240	189616	5.00	ng	0.00
27) Perylene-d12	25.49	264	178582	5.00	ng	0.00
System Monitoring Compounds						
3) 2-Fluorophenol	6.72	112	1923	0.10	ng	0.00
4) Phenol-d6	8.83	99	2387	0.10	ng	0.00
8) Nitrobenzene-d5	10.79	82	1706	0.10	ng	0.00
14) 2,4,6-Tribromophenol	18.51	330	288	0.11	ng	0.00
15) 2-Fluorobiphenyl	15.06	172	2684	0.10	ng	0.00
22) Terphenyl-d14	22.43	244	2451	0.11	ng	0.00
Target Compounds						Qvalue
2) n-Nitrosodimethylamine	3.61	42	712	0.07	ng	# 73
5) Phenol	8.86	94	2681m	0.11	ng	
6) bis(2-Chloroethyl)ether	9.01	93	1959	0.11	ng	# 98
9) Nitrobenzene	10.83	77	1939	0.10	ng	98
10) 2,4-Dimethylphenol	11.77	122	1430	0.10	ng	96
11) 2,4-Dichlorophenol	12.15	162	1049	0.10	ng	96
12) Hexachlorobutadiene	12.81	225	733	0.10	ng	96
18) Hexachlorobenzene	19.17	284	1085	0.11	ng	93
21) Benzidine	22.09	184	377	0.03	ng	# 68
23) Benzo(a)anthracene	23.77	228	15968	0.10	ng	97
24) 3,3'-Dichlorobenzidine	23.76	252	918	0.07	ng	# 88
25) Chrysene	23.82	228	17816	0.10	ng	98
26) Indeno(1,2,3-cd)pyrene	26.96	276	1863	0.03	ng	# 81
28) Benzo(b)fluoranthene	25.04	252	1592	0.07	ng	# 81
29) Benzo(k)fluoranthene	25.07	252	3442m	0.08	ng	
30) Benzo(a)pyrene	25.42	252	1683	0.06	ng	# 83
31) Dibenzo(a,h)anthracene	26.99	278	1306	0.04	ng	# 74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
Data File : BE089472.D
Acq On : 3 Feb 2015 20:15
Operator : TP/IZ
Sample : SSTDICC0.1
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

apatel
2/4/2015 2:55:13 PM

Quant Time: Feb 04 00:53:56 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 04 00:22:17 2015
Response via : Initial Calibration

