

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089473.D
 Acq On : 3 Feb 2015 20:54
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE020315

Quant Time: Feb 04 01:08:08 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00
2	n-Nitrosodimethylamine	1.000	0.960	4.0	101	0.00
3 S	2-Fluorophenol	1.000	0.967	3.3	101	0.00
4 S	Phenol-d6	1.000	0.963	3.7	99	0.00
5 C	Phenol	1.000	0.953	4.7	98	0.00
6	bis(2-Chloroethyl)ether	1.000	0.960	4.0	100	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	101	0.00
8 S	Nitrobenzene-d5	1.000	0.967	3.3	101	0.00
9	Nitrobenzene	1.000	0.963	3.7	99	0.00
10	2,4-Dimethylphenol	1.000	0.957	4.3	98	0.00
11 C	2,4-Dichlorophenol	1.000	0.934	6.6	96	0.00
12	Hexachlorobutadiene	1.000	0.966	3.4	99	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
14 S	2,4,6-Tribromophenol	1.000	0.858	14.2	90	0.00
15 S	2-Fluorobiphenyl	1.000	0.977	2.3	100	0.00
16 P	2,4-Dinitrophenol	1.000	0.937	6.3	105	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	100	0.00
18	Hexachlorobenzene	1.000	0.978	2.2	100	0.00
19 C	Pentachlorophenol	1.000	0.810	19.0	81	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	98	0.00
21	Benzidine	1.000	1.042	-4.2	99	0.00
22 S	Terphenyl-d14	1.000	0.963	3.7	95	0.00
23	Benzo(a)anthracene	1.000	0.889	11.1	86	0.00
24	3,3'-Dichlorobenzidine	1.000	0.839	16.1	84	0.00
25	Chrysene	1.000	1.025	-2.5	103	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.799	20.1	79	0.00
27 I	Perylene-d12	5.000	5.000	0.0	95	0.00
28	Benzo(b)fluoranthene	1.000	0.723	27.7#	70	0.00
29	Benzo(k)fluoranthene	1.000	1.083	-8.3	96	0.00
30 C	Benzo(a)pyrene	1.000	0.777	22.3#	74	0.00
31	Dibenzo(a,h)anthracene	1.000	0.768	23.2	72	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1