

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089480.D
 Acq On : 4 Feb 2015 1:25
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 03:04:16 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2	n-Nitrosodimethylamine	0.881	0.886	-0.6	94	0.00
3 S	2-Fluorophenol	1.617	1.551	4.1	95	0.00
4 S	Phenol-d6	2.029	1.969	3.0	95	0.00
5 C	Phenol	2.219	2.151	3.1	94	0.00
6	bis(2-Chloroethyl)ether	1.607	1.554	3.3	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.378	0.369	2.4	97	0.00
9	Nitrobenzene	0.433	0.427	1.4	97	0.00
10	2,4-Dimethylphenol	0.313	0.307	1.9	96	0.00
11 C	2,4-Dichlorophenol	0.231	0.226	2.2	96	0.00
12	Hexachlorobutadiene	0.155	0.155	0.0	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.172	0.156	9.3	92	0.00
15 S	2-Fluorobiphenyl	1.317	1.316	0.1	100	0.00
16 P	2,4-Dinitrophenol	0.037	0.012#	67.6#	69	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	97	0.00
18	Hexachlorobenzene	0.259	0.261	-0.8	100	0.00
19 C	Pentachlorophenol	0.043	0.028	34.9#	87	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
21	Benzidine	0.201	0.411	-104.5#	160#	0.00
22 S	Terphenyl-d14	0.709	0.747	-5.4	98	0.00
23	Benzo(a)anthracene	4.441	4.581	-3.2	94	0.00
24	3,3'-Dichlorobenzidine	0.346	0.393	-13.6	102	0.00
25	Chrysene	4.628	4.559	1.5	94	0.00
26	Indeno(1,2,3-cd)pyrene	1.009	1.202	-19.1	102	0.00
27 I	Perylene-d12	1.000	1.000	0.0	90	0.00
28	Benzo(b)fluoranthene	0.830	1.006	-21.2	105	0.00
29	Benzo(k)fluoranthene	1.296	1.332	-2.8	86	0.00
30 C	Benzo(a)pyrene	0.842	0.964	-14.5	98	0.00
31	Dibenzo(a,h)anthracene	0.886	1.082	-22.1	99	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1