

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089482.D
 Acq On : 4 Feb 2015 3:25
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 10:51:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	n-Nitrosodimethylamine	0.881	0.900	-2.2	86	0.00
3 S	2-Fluorophenol	1.617	1.529	5.4	84	0.00
4 S	Phenol-d6	2.029	1.976	2.6	85	0.00
5 C	Phenol	2.219	2.146	3.3	84	0.00
6	bis(2-Chloroethyl)ether	1.607	1.556	3.2	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.378	0.360	4.8	86	0.00
9	Nitrobenzene	0.433	0.422	2.5	87	0.00
10	2,4-Dimethylphenol	0.313	0.298	4.8	85	0.01
11 C	2,4-Dichlorophenol	0.231	0.220	4.8	85	0.01
12	Hexachlorobutadiene	0.155	0.154	0.6	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.172	0.144	16.3	76	0.00
15 S	2-Fluorobiphenyl	1.317	1.307	0.8	89	0.00
16 P	2,4-Dinitrophenol	0.037	0.017#	54.1#	90	0.00
17 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
18	Hexachlorobenzene	0.259	0.258	0.4	89	0.00
19 C	Pentachlorophenol	0.043	0.025	41.9#	70	0.00
20 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
21	Benzidine	0.201	0.516	-156.7#	179#	0.00
22 S	Terphenyl-d14	0.709	0.723	-2.0	85	0.00
23	Benzo(a)anthracene	4.441	4.380	1.4	81	0.00
24	3,3'-Dichlorobenzidine	0.346	0.348	-0.6	81	0.00
25	Chrysene	4.628	4.606	0.5	85	0.00
26	Indeno(1,2,3-cd)pyrene	1.009	1.036	-2.7	78	0.00
27 I	Perylene-d12	1.000	1.000	0.0	79	0.00
28	Benzo(b)fluoranthene	0.830	0.774	6.7	71	0.00
29	Benzo(k)fluoranthene	1.296	1.434	-10.6	81	0.00
30 C	Benzo(a)pyrene	0.842	0.842	0.0	75	0.00
31	Dibenzo(a,h)anthracene	0.886	0.883	0.3	71	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1