

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089482.D
 Acq On : 4 Feb 2015 3:25
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 10:51:24 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	86	0.00
2	n-Nitrosodimethylamine	1.000	0.961	3.9	86	0.00
3 S	2-Fluorophenol	1.000	0.946	5.4	84	0.00
4 S	Phenol-d6	1.000	0.974	2.6	85	0.00
5 C	Phenol	1.000	0.967	3.3	84	0.00
6	bis(2-Chloroethyl)ether	1.000	0.969	3.1	85	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	88	0.00
8 S	Nitrobenzene-d5	1.000	0.952	4.8	86	0.00
9	Nitrobenzene	1.000	0.975	2.5	87	0.00
10	2,4-Dimethylphenol	1.000	0.954	4.6	85	0.01
11 C	2,4-Dichlorophenol	1.000	0.954	4.6	85	0.01
12	Hexachlorobutadiene	1.000	0.994	0.6	88	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	89	0.00
14 S	2,4,6-Tribromophenol	1.000	0.830	17.0	76	0.00
15 S	2-Fluorobiphenyl	1.000	0.992	0.8	89	0.00
16 P	2,4-Dinitrophenol	1.000	0.924	7.6	90	0.00
17 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
18	Hexachlorobenzene	1.000	0.997	0.3	89	0.00
19 C	Pentachlorophenol	1.000	0.808	19.2	70	0.00
20 I	Chrysene-d12	5.000	5.000	0.0	83	0.00
21	Benzidine	1.000	2.146	-114.6#	179	0.00
22 S	Terphenyl-d14	1.000	1.020	-2.0	85	0.00
23	Benzo(a)anthracene	1.000	0.986	1.4	81	0.00
24	3,3'-Dichlorobenzidine	1.000	0.935	6.5	81	0.00
25	Chrysene	1.000	0.995	0.5	85	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	0.913	8.7	78	0.00
27 I	Perylene-d12	5.000	5.000	0.0	79	0.00
28	Benzo(b)fluoranthene	1.000	0.841	15.9	71	0.00
29	Benzo(k)fluoranthene	1.000	1.107	-10.7	81	0.00
30 C	Benzo(a)pyrene	1.000	0.908	9.2	75	0.00
31	Dibenzo(a,h)anthracene	1.000	0.881	11.9	71	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0