

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089490.D
 Acq On : 4 Feb 2015 10:07
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 10:52:41 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
2	n-Nitrosodimethylamine	0.881	0.890	-1.0	95	0.02
3 S	2-Fluorophenol	1.617	1.568	3.0	97	0.04
4 S	Phenol-d6	2.029	1.976	2.6	96	0.05
5 C	Phenol	2.219	1.941	12.5	86	0.05
6	bis(2-Chloroethyl)ether	1.607	1.535	4.5	95	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	96	0.00
8 S	Nitrobenzene-d5	0.378	0.382	-1.1	100	0.02
9	Nitrobenzene	0.433	0.422	2.5	96	0.01
10	2,4-Dimethylphenol	0.313	0.306	2.2	96	0.03
11 C	2,4-Dichlorophenol	0.231	0.232	-0.4	99	0.03
12	Hexachlorobutadiene	0.155	0.155	0.0	98	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	0.172	0.171	0.6	99	0.02
15 S	2-Fluorobiphenyl	1.317	1.383	-5.0	103	0.00
16 P	2,4-Dinitrophenol	0.037	0.011#	70.3#	64	0.03
17 I	Phenanthrene-d10	1.000	1.000	0.0	94	0.00
18	Hexachlorobenzene	0.259	0.259	0.0	97	0.00
19 C	Pentachlorophenol	0.043	0.019	55.8#	59	0.02
20 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
21	Benzidine	0.201	0.185	8.0	68	0.01
22 S	Terphenyl-d14	0.709	0.819	-15.5	101	0.00
23	Benzo(a)anthracene	4.441	4.664	-5.0	90	0.00
24	3,3'-Dichlorobenzidine	0.346	0.396	-14.5	97	0.00
25	Chrysene	4.628	4.558	1.5	88	0.00
26	Indeno(1,2,3-cd)pyrene	1.009	1.347	-33.5#	107	0.01
27 I	Perylene-d12	1.000	1.000	0.0	89	0.00
28	Benzo(b)fluoranthene	0.830	1.051	-26.6#	109	0.00
29	Benzo(k)fluoranthene	1.296	1.293	0.2	83	0.00
30 C	Benzo(a)pyrene	0.842	1.046	-24.2#	105	0.00
31	Dibenzo(a,h)anthracene	0.886	1.166	-31.6#	106	0.01

(#) = Out of Range

SPCC's out = 1 CCC's out = 2