

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020315\
 Data File : BE089490.D
 Acq On : 4 Feb 2015 10:07
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 10:52:41 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	97	0.00
2	n-Nitrosodimethylamine	1.000	0.952	4.8	95	0.02
3 S	2-Fluorophenol	1.000	0.970	3.0	97	0.04
4 S	Phenol-d6	1.000	0.974	2.6	96	0.05
5 C	Phenol	1.000	0.875	12.5	86	0.05
6	bis(2-Chloroethyl)ether	1.000	0.956	4.4	95	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	96	0.00
8 S	Nitrobenzene-d5	1.000	1.010	-1.0	100	0.02
9	Nitrobenzene	1.000	0.976	2.4	96	0.01
10	2,4-Dimethylphenol	1.000	0.979	2.1	96	0.03
11 C	2,4-Dichlorophenol	1.000	1.003	-0.3	99	0.03
12	Hexachlorobutadiene	1.000	1.005	-0.5	98	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	97	0.00
14 S	2,4,6-Tribromophenol	1.000	0.951	4.9	99	0.02
15 S	2-Fluorobiphenyl	1.000	1.049	-4.9	103	0.00
16 P	2,4-Dinitrophenol	1.000	0.753	24.7	64	0.03
17 I	Phenanthrene-d10	5.000	5.000	0.0	94	0.00
18	Hexachlorobenzene	1.000	1.001	-0.1	97	0.00
19 C	Pentachlorophenol	1.000	0.709	29.1#	59	0.02
20 I	Chrysene-d12	5.000	5.000	0.0	87	0.00
21	Benzidine	1.000	0.822	17.8	68	0.01
22 S	Terphenyl-d14	1.000	1.156	-15.6	101	0.00
23	Benzo(a)anthracene	1.000	1.050	-5.0	90	0.00
24	3,3'-Dichlorobenzidine	1.000	1.048	-4.8	97	0.00
25	Chrysene	1.000	0.985	1.5	88	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.147	-14.7	107	0.01
27 I	Perylene-d12	5.000	5.000	0.0	89	0.00
28	Benzo(b)fluoranthene	1.000	1.078	-7.8	109	0.00
29	Benzo(k)fluoranthene	1.000	0.997	0.3	83	0.00
30 C	Benzo(a)pyrene	1.000	1.087	-8.7	105	0.00
31	Dibenzo(a,h)anthracene	1.000	1.115	-11.5	106	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1