

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020318\
 Data File : BE095461.D
 Acq On : 4 Feb 2018 5:17
 Operator : SJ/JU
 Sample : J1314-01
 Misc :
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Feb 05 04:54:12 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 04:46:10 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1763	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	5333	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	3645	0.40	ng/ul	0.00
10) Phenanthrene-d10	0.00	188	0	0.00	ng/ul	-17.71
16) Chrysene-d12	21.80	240	7437	0.40	ng/ul	0.00
20) Perylene-d12	24.27	264	561355	0.40	ng/ul	-0.16
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	4130	0.48	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	10792	0.00	ng/ul	0.00
Target Compounds						
17) Pyrene	20.05	202	1425	0.057	ng/ul#	1




(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020318\
Data File : BE095461.D
Acq On : 4 Feb 2018 5:17
Operator : SJ/JU
Sample : J1314-01
Misc :
ALS Vial : 30 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Feb 05 04:54:12 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Feb 05 04:46:10 2018
Response via : Initial Calibration





#17

Pvrene

Concen: 0.057 ng/ul

RT: 20.05 min Scan# 2460

Delta R.T. -0.02 min

Lab File: BE095461.D

Acq: 4 Feb 2018 5:17

Instrument :

BNA_E

ClientSampleId :

Tot Ion: 202 Resp: 1425

Ion Ratio Lower Upper

202 100

101 166.6 18.2 27.4#

100 1096.8 15.6 23.4#

