

Data Path : \\74.0.250.170\svoasrv\HPCHEM1\BNA_E\Data\BE020318\
 Data File : BE095433.D
 Acq On : 3 Feb 2018 12:37
 Operator : SJ/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.438

Quant Time: Feb 05 03:30:28 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 01:47:56 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	84	0.00
3	Naphthalene	1.120	1.131	-1.0	86	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.678	-4.1	87	0.00
5	2-Methylnaphthalene	0.748	0.781	-4.4	89	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
7	Acenaphthylene	1.925	2.034	-5.7	94	0.00
8 C	Acenaphthene	1.493	1.521	-1.9	90	0.00
9	Fluorene	1.540	1.764	-14.5	101	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
11	Pentachlorophenol	0.077	0.067	13.0	80	0.00
12	Phenanthrene	1.345	1.256	6.6	80	0.00
13	Anthracene	1.231	1.154	6.3	81	0.00
14 SURR	Fluoranthene-d10	1.531	1.321	13.7	77	0.00
15 C	Fluoranthene	1.742	1.498	14.0	77	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	70	0.00
17	Pyrene	1.342	1.579	-17.7	85	0.00
18	Benzo(a)anthracene	1.251	1.332	-6.5	74	0.00
19	Chrysene	1.299	1.306	-0.5	69	0.00
20 I	Perylene-d12	1.000	1.000	0.0	74	0.00
21	Benzo(b)fluoranthene	1.411	1.389	1.6	73	0.00
22	Benzo(k)fluoranthene	1.291	1.295	-0.3	74	0.00
23 C	Benzo(a)pyrene	1.260	1.289	-2.3	76	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.373	0.7	73	0.00
25	Dibenzo(a,h)anthracene	1.130	1.090	3.5	71	0.00
26	Benzo(g,h,i)perylene	1.170	1.163	0.6	73	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0