

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020318\
 Data File : BE095448.D
 Acq On : 3 Feb 2018 21:34
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.439

Quant Time: Feb 05 03:35:26 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 03:32:57 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	82	0.00
3	Naphthalene	1.120	1.165	-4.0	85	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.702	-7.8	87	0.00
5	2-Methylnaphthalene	0.748	0.816	-9.1	90	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
7	Acenaphthylene	1.925	2.092	-8.7	95	0.00
8 C	Acenaphthene	1.493	1.523	-2.0	88	0.00
9	Fluorene	1.540	1.751	-13.7	98	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
11	Pentachlorophenol	0.077	0.063	18.2	71	0.00
12	Phenanthrene	1.345	1.293	3.9	79	0.00
13	Anthracene	1.231	1.182	4.0	80	0.00
14 SURR	Fluoranthene-d10	1.531	1.333	12.9	75	0.00
15 C	Fluoranthene	1.742	1.521	12.7	75	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	67	0.00
17	Pyrene	1.342	1.580	-17.7	81	0.00
18	Benzo(a)anthracene	1.251	1.372	-9.7	72	0.00
19	Chrysene	1.299	1.329	-2.3	67	0.00
20 I	Perylene-d12	1.000	1.000	0.0	71	0.00
21	Benzo(b)fluoranthene	1.411	1.408	0.2	71	0.00
22	Benzo(k)fluoranthene	1.291	1.279	0.9	70	0.00
23 C	Benzo(a)pyrene	1.260	1.295	-2.8	73	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.312	5.1	67	0.00
25	Dibenzo(a,h)anthracene	1.130	1.043	7.7	65	0.00
26	Benzo(g,h,i)perylene	1.170	1.111	5.0	67	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0