

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020318\
 Data File : BE095482.D
 Acq On : 4 Feb 2018 17:49
 Operator : SJ/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTD0.441

Quant Time: Feb 05 06:48:00 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Feb 05 04:46:10 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	46	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	48	0.00
3	Naphthalene	1.120	1.319	-17.8	57	0.00
4 SURR	2-Methylnaphthalene-d10	0.651	0.797	-22.4#	58	0.00
5	2-Methylnaphthalene	0.748	0.924	-23.5#	59	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	53	0.00
7	Acenaphthylene	1.925	2.307	-19.8	64	0.00
8 C	Acenaphthene	1.493	1.682	-12.7	59	0.00
9	Fluorene	1.540	1.909	-24.0#	65	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	52	0.00
11	Pentachlorophenol	0.077	0.064	16.9	45	0.00
12	Phenanthrene	1.345	1.375	-2.2	52	0.00
13	Anthracene	1.231	1.240	-0.7	52	0.00
14 SURR	Fluoranthene-d10	1.531	1.443	5.7	50	0.00
15 C	Fluoranthene	1.742	1.624	6.8	49	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	42	0.00
17	Pyrene	1.342	1.603	-19.4	52	0.00
18	Benzo(a)anthracene	1.251	1.451	-16.0	49	0.00
19	Chrysene	1.299	1.401	-7.9	45	0.00
20 I	Perylene-d12	1.000	1.000	0.0	45	0.00
21	Benzo(b)fluoranthene	1.411	1.475	-4.5	47	0.00
22	Benzo(k)fluoranthene	1.291	1.372	-6.3	47	0.00
23 C	Benzo(a)pyrene	1.260	1.357	-7.7	49	0.00
24	Indeno(1,2,3-cd)pyrene	1.382	1.411	-2.1	46	0.00
25	Dibenzo(a,h)anthracene	1.130	1.134	-0.4	45	0.00
26	Benzo(g,h,i)perylene	1.170	1.191	-1.8	45	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0