

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020415\
 Data File : BE089492.D
 Acq On : 4 Feb 2015 11:43
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 12:37:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	95	0.00
2	n-Nitrosodimethylamine	0.881	0.893	-1.4	94	0.00
3 S	2-Fluorophenol	1.617	1.540	4.8	93	0.03
4 S	Phenol-d6	2.029	1.960	3.4	93	0.04
5 C	Phenol	2.219	1.914	13.7	83	0.04
6	bis(2-Chloroethyl)ether	1.607	1.555	3.2	94	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	94	0.00
8 S	Nitrobenzene-d5	0.378	0.363	4.0	94	0.01
9	Nitrobenzene	0.433	0.425	1.8	94	0.00
10	2,4-Dimethylphenol	0.313	0.302	3.5	93	0.03
11 C	2,4-Dichlorophenol	0.231	0.229	0.9	96	0.03
12	Hexachlorobutadiene	0.155	0.154	0.6	95	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.172	0.151	12.2	85	0.01
15 S	2-Fluorobiphenyl	1.317	1.346	-2.2	97	0.00
16 P	2,4-Dinitrophenol	0.037	0.007#	81.1#	38#	0.02
17 I	Phenanthrene-d10	1.000	1.000	0.0	92	0.00
18	Hexachlorobenzene	0.259	0.262	-1.2	95	0.00
19 C	Pentachlorophenol	0.043	0.018	58.1#	53	0.01
20 I	Chrysene-d12	1.000	1.000	0.0	85	0.00
21	Benzidine	0.201	0.600	-198.5#	214#	0.00
22 S	Terphenyl-d14	0.709	0.767	-8.2	92	0.00
23	Benzo(a)anthracene	4.441	4.612	-3.9	87	0.00
24	3,3'-Dichlorobenzidine	0.346	0.401	-15.9	95	0.00
25	Chrysene	4.628	4.594	0.7	87	0.00
26	Indeno(1,2,3-cd)pyrene	1.009	1.295	-28.3#	101	0.01
27 I	Perylene-d12	1.000	1.000	0.0	88	0.00
28	Benzo(b)fluoranthene	0.830	0.993	-19.6	101	0.00
29	Benzo(k)fluoranthene	1.296	1.290	0.5	81	0.00
30 C	Benzo(a)pyrene	0.842	1.024	-21.6#	101	0.00
31	Dibenzo(a,h)anthracene	0.886	1.162	-31.2#	103	0.01

(#) = Out of Range

SPCC's out = 1 CCC's out = 2