

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020415\
 Data File : BE089492.D
 Acq On : 4 Feb 2015 11:43
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 04 12:37:01 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	95	0.00
2	n-Nitrosodimethylamine	1.000	0.955	4.5	94	0.00
3 S	2-Fluorophenol	1.000	0.953	4.7	93	0.03
4 S	Phenol-d6	1.000	0.966	3.4	93	0.04
5 C	Phenol	1.000	0.862	13.8	83	0.04
6	bis(2-Chloroethyl)ether	1.000	0.968	3.2	94	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	94	0.00
8 S	Nitrobenzene-d5	1.000	0.961	3.9	94	0.01
9	Nitrobenzene	1.000	0.981	1.9	94	0.00
10	2,4-Dimethylphenol	1.000	0.966	3.4	93	0.03
11 C	2,4-Dichlorophenol	1.000	0.992	0.8	96	0.03
12	Hexachlorobutadiene	1.000	0.995	0.5	95	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	94	0.00
14 S	2,4,6-Tribromophenol	1.000	0.863	13.7	85	0.01
15 S	2-Fluorobiphenyl	1.000	1.021	-2.1	97	0.00
16 P	2,4-Dinitrophenol	1.000	0.622	37.8#	38	0.02
17 I	Phenanthrene-d10	5.000	5.000	0.0	92	0.00
18	Hexachlorobenzene	1.000	1.011	-1.1	95	0.00
19 C	Pentachlorophenol	1.000	0.681	31.9#	53	0.01
20 I	Chrysene-d12	5.000	5.000	0.0	85	0.00
21	Benzidine	1.000	2.481	-148.1#	214	0.00
22 S	Terphenyl-d14	1.000	1.082	-8.2	92	0.00
23	Benzo(a)anthracene	1.000	1.039	-3.9	87	0.00
24	3,3'-Dichlorobenzidine	1.000	1.060	-6.0	95	0.00
25	Chrysene	1.000	0.993	0.7	87	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.108	-10.8	101	0.01
27 I	Perylene-d12	5.000	5.000	0.0	88	0.00
28	Benzo(b)fluoranthene	1.000	1.028	-2.8	101	0.00
29	Benzo(k)fluoranthene	1.000	0.995	0.5	81	0.00
30 C	Benzo(a)pyrene	1.000	1.067	-6.7	101	0.00
31	Dibenzo(a,h)anthracene	1.000	1.112	-11.2	103	0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 1