

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020415\
 Data File : BE089498.D
 Acq On : 4 Feb 2015 15:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001EC

Quant Time: Feb 04 17:30:59 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	n-Nitrosodimethylamine	0.881	0.866	1.7	81	0.00
3 S	2-Fluorophenol	1.617	1.532	5.3	83	0.02
4 S	Phenol-d6	2.029	1.949	3.9	83	0.03
5 C	Phenol	2.219	1.926	13.2	75	0.04
6	bis(2-Chloroethyl)ether	1.607	1.549	3.6	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.378	0.369	2.4	86	0.00
9	Nitrobenzene	0.433	0.422	2.5	85	0.00
10	2,4-Dimethylphenol	0.313	0.303	3.2	84	0.02
11 C	2,4-Dichlorophenol	0.231	0.225	2.6	85	0.02
12	Hexachlorobutadiene	0.155	0.153	1.3	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	0.172	0.143	16.9	72	0.01
15 S	2-Fluorobiphenyl	1.317	1.358	-3.1	89	0.00
16 P	2,4-Dinitrophenol	0.037	0.003#	91.9#	14#	0.01
17 I	Phenanthrene-d10	1.000	1.000	0.0	83	0.00
18	Hexachlorobenzene	0.259	0.261	-0.8	86	0.00
19 C	Pentachlorophenol	0.043	0.007	83.7#	18#	0.01
20 I	Chrysene-d12	1.000	1.000	0.0	77	0.00
21	Benzidine	0.201	0.306	-52.2#	99	0.00
22 S	Terphenyl-d14	0.709	0.766	-8.0	84	0.00
23	Benzo(a)anthracene	4.441	4.514	-1.6	77	0.00
24	3,3'-Dichlorobenzidine	0.346	0.369	-6.6	80	0.00
25	Chrysene	4.628	4.605	0.5	79	0.00
26	Indeno(1,2,3-cd)pyrene	1.009	1.275	-26.4#	90	0.00
27 I	Perylene-d12	1.000	1.000	0.0	79	0.00
28	Benzo(b)fluoranthene	0.830	0.976	-17.6	90	0.00
29	Benzo(k)fluoranthene	1.296	1.280	1.2	73	0.00
30 C	Benzo(a)pyrene	0.842	0.956	-13.5	85	0.00
31	Dibenzo(a,h)anthracene	0.886	1.092	-23.3	88	0.00

(#) = Out of Range

SPCC's out = 1 CCC's out = 1