

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020415\
 Data File : BE089498.D
 Acq On : 4 Feb 2015 15:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 SSTDCCC001EC

Quant Time: Feb 04 17:30:59 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIM-BE020315.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 04 00:53:32 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	85	0.00
2	n-Nitrosodimethylamine	1.000	0.928	7.2	81	0.00
3 S	2-Fluorophenol	1.000	0.948	5.2	83	0.02
4 S	Phenol-d6	1.000	0.960	4.0	83	0.03
5 C	Phenol	1.000	0.868	13.2	75	0.04
6	bis(2-Chloroethyl)ether	1.000	0.964	3.6	84	0.00
7 I	Naphthalene-d8	5.000	5.000	0.0	86	0.00
8 S	Nitrobenzene-d5	1.000	0.976	2.4	86	0.00
9	Nitrobenzene	1.000	0.975	2.5	85	0.00
10	2,4-Dimethylphenol	1.000	0.968	3.2	84	0.02
11 C	2,4-Dichlorophenol	1.000	0.975	2.5	85	0.02
12	Hexachlorobutadiene	1.000	0.991	0.9	86	0.00
13 I	Acenaphthene-d10	5.000	5.000	0.0	85	0.00
14 S	2,4,6-Tribromophenol	1.000	0.825	17.5	72	0.01
15 S	2-Fluorobiphenyl	1.000	1.031	-3.1	89	0.00
16 P	2,4-Dinitrophenol	1.000	0.489	51.1#	14	0.01
17 I	Phenanthrene-d10	5.000	5.000	0.0	83	0.00
18	Hexachlorobenzene	1.000	1.007	-0.7	86	0.00
19 C	Pentachlorophenol	1.000	0.470	53.0#	18	0.01
20 I	Chrysene-d12	5.000	5.000	0.0	77	0.00
21	Benzidine	1.000	1.307	-30.7#	99	0.00
22 S	Terphenyl-d14	1.000	1.080	-8.0	84	0.00
23	Benzo(a)anthracene	1.000	1.016	-1.6	77	0.00
24	3,3'-Dichlorobenzidine	1.000	0.986	1.4	80	0.00
25	Chrysene	1.000	0.995	0.5	79	0.00
26	Indeno(1,2,3-cd)pyrene	1.000	1.093	-9.3	90	0.00
27 I	Perylene-d12	5.000	5.000	0.0	79	0.00
28	Benzo(b)fluoranthene	1.000	1.014	-1.4	90	0.00
29	Benzo(k)fluoranthene	1.000	0.987	1.3	73	0.00
30 C	Benzo(a)pyrene	1.000	1.008	-0.8	85	0.00
31	Dibenzo(a,h)anthracene	1.000	1.054	-5.4	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1