

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020416\
 Data File : BE091386.D
 Acq On : 4 Feb 2016 9:03
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 05 06:32:04 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2 S	2-Fluorophenol	0.896	0.924	-3.1	137	0.00
3 S	Phenol-d6	1.352	1.394	-3.1	126	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	117	0.00
5 S	Nitrobenzene-d5	0.241	0.223	7.5	124	0.00
6	Naphthalene	1.128	1.022	9.4	120	0.00
7	2-Methylnaphthalene	0.609	0.581	4.6	126	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	115	0.00
9 S	2,4,6-Tribromophenol	0.153	0.131	14.4	124	0.00
10 S	2-Fluorobiphenyl	1.301	1.211	6.9	120	0.00
11	Acenaphthylene	7.133	6.639	6.9	125	0.00
12	Acenaphthene	1.373	1.224	10.9	118	0.00
13	Fluorene	1.772	1.648	7.0	119	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	114	0.00
15	Phenanthrene	1.313	1.208	8.0	119	0.00
16	Anthracene	1.100	1.032	6.2	119	0.00
17	Fluoranthene	1.203	1.127	6.3	123	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	117	0.00
19	Pyrene	1.083	1.039	4.1	126	0.00
20 S	Terphenyl-d14	0.740	0.730	1.4	126	0.00
21	Benzo(a)anthracene	0.926	0.853	7.9	127	0.00
22	Chrysene	1.371	1.247	9.0	121	0.00
23	Indeno(1,2,3-cd)pyrene	0.929	0.883	5.0	121	0.00
24 I	Perylene-d12	1.000	1.000	0.0	116	0.00
25	Benzo(b)fluoranthene	1.048	1.025	2.2	123	0.00
26	Benzo(k)fluoranthene	1.187	1.092	8.0	119	0.00
27 C	Benzo(a)pyrene	0.939	0.860	8.4	117	0.00
28	Dibenzo(a,h)anthracene	0.827	0.806	2.5	124	0.00
29	Benzo(g,h,i)perylene	1.055	1.004	4.8	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0