

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020416\
 Data File : BE091386.D
 Acq On : 4 Feb 2016 9:03
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 05 06:32:04 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	116	0.00
2 S	2-Fluorophenol	1.000	0.979	2.1	137	0.00
3 S	Phenol-d6	1.000	1.015	-1.5	126	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	117	0.00
5 S	Nitrobenzene-d5	1.000	0.915	8.5	124	0.00
6	Naphthalene	1.000	0.907	9.3	120	0.00
7	2-Methylnaphthalene	1.000	0.935	6.5	126	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	115	0.00
9 S	2,4,6-Tribromophenol	1.000	0.851	14.9	124	0.00
10 S	2-Fluorobiphenyl	1.000	0.930	7.0	120	0.00
11	Acenaphthylene	1.000	0.913	8.7	125	0.00
12	Acenaphthene	1.000	0.892	10.8	118	0.00
13	Fluorene	1.000	0.930	7.0	119	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	114	0.00
15	Phenanthrene	1.000	0.920	8.0	119	0.00
16	Anthracene	1.000	0.938	6.2	119	0.00
17	Fluoranthene	1.000	0.917	8.3	123	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	117	0.00
19	Pyrene	1.000	0.959	4.1	126	0.00
20 S	Terphenyl-d14	1.000	0.987	1.3	126	0.00
21	Benzo(a)anthracene	1.000	0.910	9.0	127	0.00
22	Chrysene	1.000	0.949	5.1	121	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	0.913	8.7	121	0.00
24 I	Perylene-d12	5.000	5.000	0.0	116	0.00
25	Benzo(b)fluoranthene	1.000	0.947	5.3	123	0.00
26	Benzo(k)fluoranthene	1.000	0.889	11.1	119	0.00
27 C	Benzo(a)pyrene	1.000	0.893	10.7	117	0.00
28	Dibenzo(a,h)anthracene	1.000	0.922	7.8	124	0.00
29	Benzo(g,h,i)perylene	1.000	0.928	7.2	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0