

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020416\
 Data File : BE091393.D
 Acq On : 4 Feb 2016 17:15
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 05 06:33:40 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	102	0.00
2 S	2-Fluorophenol	1.000	0.942	5.8	114	0.00
3 S	Phenol-d6	1.000	0.991	0.9	108	0.00
4 I	Naphthalene-d8	5.000	5.000	0.0	104	0.00
5 S	Nitrobenzene-d5	1.000	0.954	4.6	117	0.00
6	Naphthalene	1.000	0.918	8.2	107	0.00
7	2-Methylnaphthalene	1.000	0.949	5.1	113	0.00
8 I	Acenaphthene-d10	5.000	5.000	0.0	101	0.00
9 S	2,4,6-Tribromophenol	1.000	0.844	15.6	108	0.00
10 S	2-Fluorobiphenyl	1.000	0.952	4.8	108	0.00
11	Acenaphthylene	1.000	0.917	8.3	111	0.00
12	Acenaphthene	1.000	0.887	11.3	103	0.00
13	Fluorene	1.000	0.918	8.2	103	0.00
14 I	Phenanthrene-d10	5.000	5.000	0.0	101	0.00
15	Phenanthrene	1.000	0.910	9.0	104	0.00
16	Anthracene	1.000	0.941	5.9	105	0.00
17	Fluoranthene	1.000	0.915	8.5	109	0.00
18 I	Chrysene-d12	5.000	5.000	0.0	104	0.00
19	Pyrene	1.000	0.941	5.9	110	0.00
20 S	Terphenyl-d14	1.000	0.975	2.5	111	0.00
21	Benzo(a)anthracene	1.000	0.894	10.6	111	0.00
22	Chrysene	1.000	0.945	5.5	108	0.00
23	Indeno(1,2,3-cd)pyrene	1.000	0.916	8.4	108	0.00
24 I	Perylene-d12	5.000	5.000	0.0	103	0.00
25	Benzo(b)fluoranthene	1.000	0.917	8.3	105	0.00
26	Benzo(k)fluoranthene	1.000	0.922	7.8	110	0.00
27 C	Benzo(a)pyrene	1.000	0.897	10.3	105	0.00
28	Dibenzo(a,h)anthracene	1.000	0.926	7.4	112	0.00
29	Benzo(g,h,i)perylene	1.000	0.927	7.3	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0