

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089501.D
 Acq On : 5 Feb 2015 12:01
 Operator : TP/IZ
 Sample : SSTDICC005
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC005

Quant Time: Feb 06 02:48:23 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 02:40:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	28909	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	120078	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	57584	5.00	ng	0.00
11) Phenanthrene-d10	19.90	188	110666	5.00	ng	0.00
16) Chrysene-d12	23.77	240	120459	5.00	ng	0.00
22) Perylene-d12	25.46	264	117266	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	47256	5.16	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	84171	4.64	ng	0.00
18) Terphenyl-d14	22.40	244	101687	4.85	ng	0.00

Target Compounds				Ovalue		
4) Naphthalene	12.45	128	143543	4.75	ng	99
5) 2-Methylnaphthalene	14.11	142	95923	4.84	ng	100
8) Acenaphthylene	16.24	152	611313	4.95	ng	100
9) Acenaphthene	16.66	154	88061	4.64	ng	99
10) Fluorene	17.91	166	102926	4.87	ng	99
12) Pentachlorophenol	19.58	266	7245	6.66	ng	94
13) Phenanthrene	19.95	178	156838	4.67	ng	99
14) Anthracene	20.04	178	153262	5.10	ng	100
15) Fluoranthene	21.84	202	160761	5.01	ng	100
17) Pyrene	22.15	202	167759	4.87	ng	100
19) Benzo(a)anthracene	23.75	228	637936	4.83	ng	97
20) Chrysene	23.79	228	630076	4.64	ng	97
21) Indeno(1,2,3-cd)pyrene	26.93	276	727501	5.15	ng	99
23) Benzo(b)fluoranthene	25.02	252	150749	5.53	ng	99
24) Benzo(k)fluoranthene	25.05	252	177270	4.74	ng	99
25) Benzo(a)pyrene	25.39	252	152068	5.50	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	156043	5.29	ng	99
27) Benzo(a,h,i)perylene	27.36	276	592779	4.92	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089501.D
Acq On : 5 Feb 2015 12:01
Operator : TP/IZ
Sample : SSTDICC005
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC005

Quant Time: Feb 06 02:48:23 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 02:40:17 2015
Response via : Initial Calibration

