

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089503.D
 Acq On : 5 Feb 2015 13:18
 Operator : TP/IZ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICCC001

Quant Time: Feb 06 02:49:07 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 02:40:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	31396	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	132520	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	65560	5.00	ng	0.00
11) Phenanthrene-d10	19.90	188	128634	5.00	ng	0.00
16) Chrysene-d12	23.76	240	143431	5.00	ng	0.00
22) Perylene-d12	25.46	264	145552	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	10361	1.02	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	23554	1.14	ng	0.00
18) Terphenyl-d14	22.40	244	29874	1.20	ng	0.00

Target Compounds					Ovalue	
4) Naphthalene	12.45	128	34641	1.04	ng	100
5) 2-Methylnaphthalene	14.11	142	25028	1.15	ng	100
8) Acenaphthylene	16.24	152	166466	1.18	ng	100
9) Acenaphthene	16.65	154	25298	1.17	ng	100
10) Fluorene	17.90	166	29308	1.22	ng	100
12) Pentachlorophenol	19.58	266	317	0.25	ng	100
13) Phenanthrene	19.94	178	46246	1.18	ng	100
14) Anthracene	20.04	178	42845	1.23	ng	100
15) Fluoranthene	21.84	202	46146	1.24	ng	100
17) Pyrene	22.15	202	49120	1.20	ng	100
19) Benzo(a)anthracene	23.75	228	190352	1.21	ng	100
20) Chrysene	23.79	228	193892	1.20	ng	100
21) Indeno(1,2,3-cd)pyrene	26.93	276	223160	1.33	ng	100
23) Benzo(b)fluoranthene	25.02	252	41159	1.22	ng	100
24) Benzo(k)fluoranthene	25.05	252	55685	1.20	ng	100
25) Benzo(a)pyrene	25.39	252	42142	1.23	ng	100
26) Dibenzo(a,h)anthracene	26.95	278	47343	1.29	ng	100
27) Benzo(a,h,i)perylene	27.36	276	186516	1.25	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089503.D
Acq On : 5 Feb 2015 13:18
Operator : TP/IZ
Sample : SSTDICCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICCC001

Quant Time: Feb 06 02:49:07 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 02:40:17 2015
Response via : Initial Calibration

