

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089504.D
 Acq On : 5 Feb 2015 14:06
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Feb 06 02:49:31 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 02:40:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	29321	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	117123	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	51363	5.00	ng	0.00
11) Phenanthrene-d10	19.90	188	97991	5.00	ng	0.00
16) Chrysene-d12	23.76	240	107509	5.00	ng	0.00
22) Perylene-d12	25.46	264	107750	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	6395	0.72	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	11151	0.69	ng	0.00
18) Terphenyl-d14	22.40	244	12590	0.67	ng	0.00
Target Compounds						Ovalue
4) Naphthalene	12.45	128	21152	0.72	ng	100
5) 2-Methylnaphthalene	14.10	142	12839	0.66	ng	95
8) Acenaphthylene	16.24	152	75586	0.69	ng	100
9) Acenaphthene	16.65	154	11408	0.67	ng	98
10) Fluorene	17.91	166	12658	0.67	ng	98
12) Pentachlorophenol	19.60	266	37	0.04	ng	# 85
13) Phenanthrene	19.94	178	20368	0.68	ng	100
14) Anthracene	20.04	178	18318	0.69	ng	100
15) Fluoranthene	21.84	202	19252	0.68	ng	99
17) Pyrene	22.15	202	20661	0.67	ng	99
19) Benzo(a)anthracene	23.75	228	79161	0.67	ng	99
20) Chrysene	23.79	228	82015	0.68	ng	99
21) Indeno(1,2,3-cd)pyrene	26.92	276	88359	0.70	ng	98
23) Benzo(b)fluoranthene	25.02	252	15828	0.63	ng	98
24) Benzo(k)fluoranthene	25.05	252	23705	0.69	ng	98
25) Benzo(a)pyrene	25.39	252	16371	0.64	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	18536	0.68	ng	96
27) Benzo(a,h,i)perylene	27.36	276	75204	0.68	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089504.D
Acq On : 5 Feb 2015 14:06
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Feb 06 02:49:31 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 02:40:17 2015
Response via : Initial Calibration

