

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089505.D
 Acq On : 5 Feb 2015 14:45
 Operator : TP/IZ
 Sample : SSTDICC0.5
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.5

Quant Time: Feb 06 03:14:27 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 02:40:17 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	27470	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	119764	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	55756	5.00	ng	0.00
11) Phenanthrene-d10	19.91	188	108190	5.00	ng	0.00
16) Chrysene-d12	23.76	240	117419	5.00	ng	0.00
22) Perylene-d12	25.46	264	115596	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	4241	0.46	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	8376	0.48	ng	0.00
18) Terphenyl-d14	22.40	244	9507	0.47	ng	0.00
Target Compounds						Ovalue
4) Naphthalene	12.44	128	14976	0.50	ng	99
5) 2-Methylnaphthalene	14.10	142	9503	0.48	ng	95
8) Acenaphthylene	16.24	152	55169	0.46	ng	100
9) Acenaphthene	16.65	154	8611	0.47	ng	99
10) Fluorene	17.91	166	9570	0.47	ng	98
13) Phenanthrene	19.94	178	15517	0.47	ng	99
14) Anthracene	20.03	178	13589	0.46	ng	99
15) Fluoranthene	21.83	202	14511	0.46	ng	100
17) Pyrene	22.15	202	15479	0.46	ng	100
19) Benzo(a)anthracene	23.75	228	59163	0.46	ng	100
20) Chrysene	23.79	228	62712	0.47	ng	100
21) Indeno(1,2,3-cd)pyrene	26.92	276	63831	0.46	ng	97
23) Benzo(b)fluoranthene	25.02	252	11598	0.43	ng	97
24) Benzo(k)fluoranthene	25.05	252	17436	0.47	ng	97
25) Benzo(a)pyrene	25.39	252	11913	0.44	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	13182	0.45	ng	97
27) Benzo(g,h,i)perylene	27.36	276	54772	0.46	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089505.D
Acq On : 5 Feb 2015 14:45
Operator : TP/IZ
Sample : SSTDICC0.5
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.5

Quant Time: Feb 06 03:14:27 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 02:40:17 2015
Response via : Initial Calibration

