

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089508.D
 Acq On : 5 Feb 2015 18:16
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICVBE020515

Quant Time: Feb 06 12:51:13 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	30185	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	129266	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	63342	5.00	ng	0.00
11) Phenanthrene-d10	19.91	188	122482	5.00	ng	0.00
16) Chrysene-d12	23.76	240	136802	5.00	ng	0.00
22) Perylene-d12	25.46	264	136624	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	10766	1.09	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	24297	1.22	ng	0.00
18) Terphenyl-d14	22.41	244	30027	1.26	ng	0.00
Target Compounds						Ovalue
4) Naphthalene	12.45	128	36800	1.13	ng	99
5) 2-Methylnaphthalene	14.10	142	26323	1.23	ng	96
8) Acenaphthylene	16.24	152	168225	1.24	ng	100
9) Acenaphthene	16.65	154	25203	1.21	ng	99
10) Fluorene	17.91	166	29140	1.12	ng	99
12) Pentachlorophenol	19.58	266	74	0.06	ng	# 80
13) Phenanthrene	19.94	178	46257	1.24	ng	100
14) Anthracene	20.03	178	42407	1.12	ng	100
15) Fluoranthene	21.83	202	45831	1.13	ng	99
17) Pyrene	22.15	202	48907	1.25	ng	99
19) Benzo(a)anthracene	23.75	228	187758	1.13	ng	99
20) Chrysene	23.79	228	193282	1.16	ng	99
21) Indeno(1,2,3-cd)pyrene	26.92	276	217119	1.14	ng	97
23) Benzo(b)fluoranthene	25.02	252	41766	1.10	ng	98
24) Benzo(k)fluoranthene	25.05	252	55958	1.28	ng	98
25) Benzo(a)pyrene	25.39	252	40677	1.09	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	46208	1.14	ng	100
27) Benzo(a,h,i)perylene	27.36	276	182417	1.16	ng	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089508.D
Acq On : 5 Feb 2015 18:16
Operator : TP/IZ
Sample : SSTDICV001
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
ICVBE020515

Quant Time: Feb 06 12:51:13 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 12:30:39 2015
Response via : Initial Calibration

