

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089509.D
 Acq On : 5 Feb 2015 18:55
 Operator : TP/IZ
 Sample : PB81495BS
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB81495BS

Quant Time: Feb 06 04:01:36 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 03:21:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	27059	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	118419	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	58451	5.00	ng	0.00
11) Phenanthrene-d10	19.90	188	111968	5.00	ng	0.00
16) Chrysene-d12	23.76	240	120487	5.00	ng	0.00
22) Perylene-d12	25.46	264	115623	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	52736	5.84	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	95926	5.21	ng	0.00
18) Terphenyl-d14	22.41	244	119165	5.68	ng	0.00
Target Compounds						Ovalue
4) Naphthalene	12.45	128	146714	4.92	ng	99
5) 2-Methylnaphthalene	14.10	142	95583	4.89	ng	94
8) Acenaphthylene	16.24	152	616321	4.91	ng	100
9) Acenaphthene	16.65	154	91015	4.73	ng	99
10) Fluorene	17.91	166	108948	5.08	ng	99
12) Pentachlorophenol	19.58	266	13079	0.24	ng	94
13) Phenanthrene	19.94	178	168088	4.94	ng	99
14) Anthracene	20.04	178	168096	5.38	ng	100
15) Fluoranthene	21.83	202	171637	5.21	ng	99
17) Pyrene	22.15	202	184279	5.35	ng	99
19) Benzo(a)anthracene	23.75	228	687615	5.30	ng	99
20) Chrysene	23.79	228	690506	5.39	ng	99
21) Indeno(1,2,3-cd)pyrene	26.92	276	741647	4.95	ng	98
23) Benzo(b)fluoranthene	25.02	252	153282	5.07	ng	99
24) Benzo(k)fluoranthene	25.05	252	196519	5.33	ng	97
25) Benzo(a)pyrene	25.39	252	157230	5.22	ng	98
26) Dibenzo(a,h)anthracene	26.95	278	164234	5.23	ng	98
27) Benzo(a,h,i)perylene	27.36	276	626530	5.24	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089509.D
Acq On : 5 Feb 2015 18:55
Operator : TP/IZ
Sample : PB81495BS
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB81495BS

Quant Time: Feb 06 04:01:36 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 03:21:38 2015
Response via : Initial Calibration

