

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089513.D
 Acq On : 5 Feb 2015 21:30
 Operator : TP/IZ
 Sample : PB81591BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB81591BS

Quant Time: Feb 06 04:12:21 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 03:21:38 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	24116	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	97889	5.00	ng	0.00
6) Acenaphthene-d10	16.59	164	43971	5.00	ng	0.00
11) Phenanthrene-d10	19.90	188	84532	5.00	ng	0.00
16) Chrysene-d12	23.76	240	93979	5.00	ng	0.00
22) Perylene-d12	25.46	264	89956	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.76	82	53912	7.22	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	96154	6.94	ng	0.00
18) Terphenyl-d14	22.40	244	125751	7.69	ng	0.00

Target Compounds				Ovalue		
4) Naphthalene	12.44	128	127655	5.18	ng	99
5) 2-Methylnaphthalene	14.10	142	81744	5.06	ng	96
8) Acenaphthylene	16.24	152	525899	5.57	ng	100
9) Acenaphthene	16.65	154	76930	5.31	ng	99
10) Fluorene	17.91	166	94899	6.11	ng	99
12) Pentachlorophenol	19.58	266	10938	0.27	ng	95
13) Phenanthrene	19.94	178	147399	5.74	ng	100
14) Anthracene	20.04	178	148222	6.55	ng	100
15) Fluoranthene	21.83	202	152041	6.42	ng	99
17) Pyrene	22.15	202	165004	6.14	ng	99
19) Benzo(a)anthracene	23.75	228	618158	6.35	ng	98
20) Chrysene	23.79	228	640389	6.74	ng	99
21) Indeno(1,2,3-cd)pyrene	26.92	276	675036	6.08	ng	98
23) Benzo(b)fluoranthene	25.02	252	140676	6.23	ng	99
24) Benzo(k)fluoranthene	25.05	252	178367	6.22	ng	97
25) Benzo(a)pyrene	25.39	252	142857	6.29	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	149773	6.38	ng	97
27) Benzo(a,h,i)perylene	27.36	276	572421	6.40	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089513.D
Acq On : 5 Feb 2015 21:30
Operator : TP/IZ
Sample : PB81591BS
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB81591BS

Quant Time: Feb 06 04:12:21 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 03:21:38 2015
Response via : Initial Calibration

