

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
 Data File : BE089517.D
 Acq On : 6 Feb 2015 1:35
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 06 13:14:33 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.49	152	25640	5.00	ng	0.00
2) Naphthalene-d8	12.40	136	104624	5.00	ng	0.00
6) Acenaphthene-d10	16.58	164	50837	5.00	ng	-0.01
11) Phenanthrene-d10	19.90	188	101399	5.00	ng	0.00
16) Chrysene-d12	23.76	240	114457	5.00	ng	0.00
22) Perylene-d12	25.46	264	111502	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.75	82	7166	0.90	ng	0.00
7) 2-Fluorobiphenyl	15.03	172	15540	0.97	ng	0.00
18) Terphenyl-d14	22.40	244	18866	0.95	ng	0.00
Target Compounds						Ovalue
4) Naphthalene	12.44	128	26081	0.99	ng	99
5) 2-Methylnaphthalene	14.10	142	17115	0.99	ng	98
8) Acenaphthylene	16.23	152	106003	0.97	ng	100
9) Acenaphthene	16.65	154	16442	0.98	ng	99
10) Fluorene	17.91	166	18638	0.90	ng	100
13) Phenanthrene	19.94	178	30696	1.00	ng	99
14) Anthracene	20.03	178	26786	0.87	ng	100
15) Fluoranthene	21.83	202	28960	0.87	ng	99
17) Pyrene	22.15	202	31833	0.97	ng	100
19) Benzo(a)anthracene	23.75	228	117911	0.85	ng	97
20) Chrysene	23.79	228	129527	0.92	ng	97
21) Indeno(1,2,3-cd)pyrene	26.92	276	127149	0.81	ng	98
23) Benzo(b)fluoranthene	25.02	252	25004	0.83	ng	99
24) Benzo(k)fluoranthene	25.04	252	35829	1.01	ng	99
25) Benzo(a)pyrene	25.39	252	24562	0.82	ng	99
26) Dibenzo(a,h)anthracene	26.95	278	26859	0.82	ng	98
27) Benzo(g,h,i)perylene	27.35	276	109536	0.85	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020515\
Data File : BE089517.D
Acq On : 6 Feb 2015 1:35
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Feb 06 13:14:33 2015
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 06 12:30:39 2015
Response via : Initial Calibration

