

Data Path : Z:\HPCHEM1\BNA E\Data\BE020515\
 Data File : BE089515.D
 Acq On : 5 Feb 2015 22:48
 Operator : TP/IZ
 Sample : SSTDCCC001EC
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001EC

Quant Time: Feb 06 04:13:03 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 03:21:38 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	74	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	74	0.00
3 S	Nitrobenzene-d5	0.381	0.516	-35.4#	98	0.00
4	Naphthalene	1.259	1.616	-28.4#	92	0.00
5	2-Methylnaphthalene	0.825	1.211	-46.8#	95	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	73	-0.01
7 S	2-Fluorobiphenyl	1.575	2.775	-76.2#	113	0.00
8	Acenaphthylene	10.727	16.891	-57.5#	97	-0.01
9	Acenaphthene	1.646	2.553	-55.1#	97	0.00
10	Fluorene	1.836	3.135	-70.8#	103	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
12	Pentachlorophenol	0.050	0.015	70.0#	89	0.00
13	Phenanthrene	1.518	2.585	-70.3#	107	0.00
14	Anthracene	1.358	2.436	-79.4#	109	0.00
15	Fluoranthene	1.450	2.577	-77.7#	107	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
17	Pyrene	1.429	2.452	-71.6#	109	0.00
18 S	Terphenyl-d14	0.870	1.731	-99.0#	127	0.00
19	Benzo(a)anthracene	5.486	9.320	-69.9#	107	0.00
20	Chrysene	5.638	10.018	-77.7#	113	0.00
21	Indeno(1,2,3-cd)pyrene	5.865	10.691	-82.3#	105	0.00
22 I	Perylene-d12	1.000	1.000	0.0	74	0.00
23	Benzo(b)fluoranthene	1.162	2.178	-87.4#	115	0.00
24	Benzo(k)fluoranthene	1.594	2.784	-74.7#	108	0.00
25 C	Benzo(a)pyrene	1.178	2.125	-80.4#	109	0.00
26	Dibenzo(a,h)anthracene	1.257	2.344	-86.5#	107	0.00
27	Benzo(a,h,i)perylene	5.133	9.240	-80.0#	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1