

Data Path : Z:\HPCHEM1\BNA E\Data\BE020515\
 Data File : BE089522.D
 Acq On : 6 Feb 2015 4:49
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc : 4 TIME INT STD
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 06 13:24:34 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	116	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	139	0.00
3 S	Nitrobenzene-d5	0.381	0.385	-1.0	137	0.00
4	Naphthalene	1.259	0.767	39.1#	82	0.00
5	2-Methylnaphthalene	0.825	0.451	45.3#	66	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	156#	0.00
7 S	2-Fluorobiphenyl	1.575	1.901	-20.7	165#	0.00
8	Acenaphthylene	10.727	4.949	53.9#	61	0.00
9	Acenaphthene	1.646	0.740	55.0#	60	0.00
10	Fluorene	1.836	0.836	54.5#	58	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	160#	0.00
12	Pentachlorophenol	0.050	0.001	98.0#	8#	0.00
13	Phenanthrene	1.518	0.668	56.0#	60	0.00
14	Anthracene	1.358	0.596	56.1#	57	0.00
15	Fluoranthene	1.450	0.651	55.1#	58	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	163#	0.00
17	Pyrene	1.429	0.606	57.6#	58	0.00
18 S	Terphenyl-d14	0.870	1.188	-36.6#	186#	0.00
19	Benzo(a)anthracene	5.486	2.386	56.5#	59	0.00
20	Chrysene	5.638	2.415	57.2#	58	0.00
21	Indeno(1,2,3-cd)pyrene	5.865	2.386	59.3#	50	0.00
22 I	Perylene-d12	1.000	1.000	0.0	155#	0.00
23	Benzo(b)fluoranthene	1.162	0.536	53.9#	59	0.00
24	Benzo(k)fluoranthene	1.594	0.686	57.0#	55	0.00
25 C	Benzo(a)pyrene	1.178	0.498	57.7#	53	0.00
26	Dibenzo(a,h)anthracene	1.257	0.529	57.9#	50	0.00
27	Benzo(a,h,i)perylene	5.133	2.082	59.4#	50	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1