

Data Path : Z:\HPCHEM1\BNA E\Data\BE020515\
 Data File : BE089522.D
 Acq On : 6 Feb 2015 4:49
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc : 4 TIME INT STD
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 06 13:24:34 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	116	0.00
2 I	Naphthalene-d8	5.000	5.000	0.0	139	0.00
3 S	Nitrobenzene-d5	1.000	1.009	-0.9	137	0.00
4	Naphthalene	1.000	0.609	39.1#	82	0.00
5	2-Methylnaphthalene	1.000	0.547	45.3#	66	0.00
6 I	Acenaphthene-d10	5.000	5.000	0.0	156	0.00
7 S	2-Fluorobiphenyl	1.000	1.207	-20.7	165	0.00
8	Acenaphthylene	1.000	0.461	53.9#	61	0.00
9	Acenaphthene	1.000	0.450	55.0#	60	0.00
10	Fluorene	1.000	0.423	57.7#	58	0.00
11 I	Phenanthrene-d10	5.000	5.000	0.0	160	0.00
12	Pentachlorophenol	1.000	0.012	98.8#	8	0.00
13	Phenanthrene	1.000	0.440	56.0#	60	0.00
14	Anthracene	1.000	0.418	58.2#	57	0.00
15	Fluoranthene	1.000	0.424	57.6#	58	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	163	0.00
17	Pyrene	1.000	0.424	57.6#	58	0.00
18 S	Terphenyl-d14	1.000	1.365	-36.5#	186	0.00
19	Benzo(a)anthracene	1.000	0.408	59.2#	59	0.00
20	Chrysene	1.000	0.395	60.5#	58	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	0.379	62.1#	50	0.00
22 I	Perylene-d12	5.000	5.000	0.0	155	0.00
23	Benzo(b)fluoranthene	1.000	0.451	54.9#	59	0.00
24	Benzo(k)fluoranthene	1.000	0.430	57.0#	55	0.00
25 C	Benzo(a)pyrene	1.000	0.425	57.5#	53	0.00
26	Dibenzo(a,h)anthracene	1.000	0.396	60.4#	50	0.00
27	Benzo(a,h,i)perylene	1.000	0.377	62.3#	50	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1