

Data Path : Z:\HPCHEM1\BNA E\Data\BE020515\
 Data File : Be089508.d
 Acq On : 5 Feb 2015 18:16
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE020515

Quant Time: Feb 06 12:51:13 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
3 S	Nitrobenzene-d5	0.381	0.416	-9.2	104	0.00
4	Naphthalene	1.259	1.423	-13.0	106	0.00
5	2-Methylnaphthalene	0.825	1.018	-23.4	105	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	97	0.00
7 S	2-Fluorobiphenyl	1.575	1.918	-21.8	103	0.00
8	Acenaphthylene	10.727	13.279	-23.8	101	0.00
9	Acenaphthene	1.646	1.989	-20.8	100	0.00
10	Fluorene	1.836	2.300	-25.3#	99	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	95	0.00
12	Pentachlorophenol	0.050	0.003	94.0#	23#	0.00
13	Phenanthrene	1.518	1.888	-24.4	100	0.00
14	Anthracene	1.358	1.731	-27.5#	99	0.00
15	Fluoranthene	1.450	1.871	-29.0#	99	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	95	0.00
17	Pyrene	1.429	1.788	-25.1#	100	0.00
18 S	Terphenyl-d14	0.870	1.097	-26.1#	101	0.00
19	Benzo(a)anthracene	5.486	6.862	-25.1#	99	0.00
20	Chrysene	5.638	7.064	-25.3#	100	0.00
21	Indeno(1,2,3-cd)pyrene	5.865	7.936	-35.3#	97	0.00
22 I	Perylene-d12	1.000	1.000	0.0	94	0.00
23	Benzo(b)fluoranthene	1.162	1.529	-31.6#	101	0.00
24	Benzo(k)fluoranthene	1.594	2.048	-28.5#	100	0.00
25 C	Benzo(a)pyrene	1.178	1.489	-26.4#	97	0.00
26	Dibenzo(a,h)anthracene	1.257	1.691	-34.5#	98	0.00
27	Benzo(a,h,i)perylene	5.133	6.676	-30.1#	98	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1