

Data Path : Z:\HPCHEM1\BNA E\Data\BE020515\
 Data File : Be089517.d
 Acq On : 6 Feb 2015 1:35
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 06 13:14:33 2015
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SIMPAH-BE020515.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 06 12:30:39 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	82	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	79	0.00
3 S	Nitrobenzene-d5	0.381	0.342	10.2	69	0.00
4	Naphthalene	1.259	1.246	1.0	75	0.00
5	2-Methylnaphthalene	0.825	0.818	0.8	68	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	78	-0.01
7 S	2-Fluorobiphenyl	1.575	1.528	3.0	66	0.00
8	Acenaphthylene	10.727	10.426	2.8	64	-0.01
9	Acenaphthene	1.646	1.617	1.8	65	0.00
10	Fluorene	1.836	1.833	0.2	64	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
12	Pentachlorophenol	0.050	0.001	98.0#	6#	0.00
13	Phenanthrene	1.518	1.514	0.3	66	0.00
14	Anthracene	1.358	1.321	2.7	63	0.00
15	Fluoranthene	1.450	1.428	1.5	63	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	80	0.00
17	Pyrene	1.429	1.391	2.7	65	0.00
18 S	Terphenyl-d14	0.870	0.824	5.3	63	0.00
19	Benzo(a)anthracene	5.486	5.151	6.1	62	0.00
20	Chrysene	5.638	5.658	-0.4	67	0.00
21	Indeno(1,2,3-cd)pyrene	5.865	5.554	5.3	57	0.00
22 I	Perylene-d12	1.000	1.000	0.0	77	0.00
23	Benzo(b)fluoranthene	1.162	1.121	3.5	61	0.00
24	Benzo(k)fluoranthene	1.594	1.607	-0.8	64	0.00
25 C	Benzo(a)pyrene	1.178	1.101	6.5	58	0.00
26	Dibenzo(a,h)anthracene	1.257	1.204	4.2	57	0.00
27	Benzo(a,h,i)perylene	5.133	4.912	4.3	59	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0