

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020516\
 Data File : BE091399.D
 Acq On : 5 Feb 2016 14:38
 Operator : UM/IZ
 Sample : PB88200BS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB88200BS

Quant Time: Feb 05 16:43:32 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	42450	5.00	ng	0.00
4) Naphthalene-d8	9.54	136	207693	5.00	ng	0.00
8) Acenaphthene-d10	13.35	164	129891	5.00	ng	0.00
14) Phenanthrene-d10	16.08	188	317838	5.00	ng	-0.01
18) Chrysene-d12	20.21	240	399790	5.00	ng	0.00
24) Perylene-d12	22.09	264	379579	5.00	ng	0.00

System Monitoring Compounds						
2) 2-Fluorophenol	4.91	112	67822	6.76	ng	-0.04
3) Phenol-d6	6.53	99	105066	7.37	ng	-0.04
5) Nitrobenzene-d5	8.14	82	52423	3.88	ng	-0.03
9) 2,4,6-Tribromophenol	14.89	330	43176	6.84	ng	-0.01
10) 2-Fluorobiphenyl	12.00	172	150396	4.45	ng	-0.01
20) Terphenyl-d14	18.61	244	251429	4.25	ng	0.00

Target Compounds				Qvalue		
6) Naphthalene	9.58	128	170359	3.64	ng	99
7) 2-Methylnaphthalene	11.11	142	114728	3.96	ng	97
11) Acenaphthylene	13.09	152	826223	3.69	ng	100
12) Acenaphthene	13.41	154	130278	3.65	ng	97
13) Fluorene	14.39	166	182731	3.97	ng	99
15) Phenanthrene	16.12	178	319440	3.83	ng	99
16) Anthracene	16.20	178	307337	4.39	ng	99
17) Fluoranthene	18.08	202	352568	4.05	ng	99
19) Pyrene	18.46	202	377555	4.36	ng	99
21) Benzo(a)anthracene	20.18	228	353693	4.00	ng	97
22) Chrysene	20.24	228	396611	3.99	ng	97
23) Indeno(1,2,3-cd)pyrene	23.69	276	354779	3.94	ng	99
25) Benzo(b)fluoranthene	21.54	252	371220	3.96	ng	98
26) Benzo(k)fluoranthene	21.57	252	398081	4.20	ng	98
27) Benzo(a)pyrene	22.00	252	340946	3.94	ng	99
28) Dibenzo(a,h)anthracene	23.64	278	317613	3.74	ng	99
29) Benzo(g,h,i)perylene	24.24	276	364919	3.90	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020516\
Data File : BE091399.D
Acq On : 5 Feb 2016 14:38
Operator : UM/IZ
Sample : PB88200BS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB88200BS

Quant Time: Feb 05 16:43:32 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 00:27:53 2016
Response via : Initial Calibration

