

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020516\
 Data File : BE091400.D
 Acq On : 5 Feb 2016 15:14
 Operator : UM/IZ
 Sample : PB88200BSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 PB88200BSD

Quant Time: Feb 05 16:45:10 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	6.87	152	38138	5.00	ng	0.00
4) Naphthalene-d8	9.55	136	183303	5.00	ng	0.00
8) Acenaphthene-d10	13.35	164	115389	5.00	ng	0.00
14) Phenanthrene-d10	16.08	188	286259	5.00	ng	-0.01
18) Chrysene-d12	20.21	240	360209	5.00	ng	0.00
24) Perylene-d12	22.09	264	339743	5.00	ng	0.00

System Monitoring Compounds						
2) 2-Fluorophenol	4.91	112	49755	5.56	ng	-0.04
3) Phenol-d6	6.54	99	75936	5.97	ng	-0.03
5) Nitrobenzene-d5	8.15	82	38544	3.28	ng	-0.02
9) 2,4,6-Tribromophenol	14.89	330	32507	5.85	ng	-0.01
10) 2-Fluorobiphenyl	12.00	172	110317	3.67	ng	-0.01
20) Terphenyl-d14	18.62	244	190171	3.57	ng	0.00

Target Compounds				Qvalue		
6) Naphthalene	9.58	128	120520	2.92	ng	100
7) 2-Methylnaphthalene	11.11	142	80027	3.16	ng	97
11) Acenaphthylene	13.10	152	591643	3.01	ng	100
12) Acenaphthene	13.41	154	93613	2.95	ng	98
13) Fluorene	14.39	166	134099	3.28	ng	100
15) Phenanthrene	16.12	178	238295	3.17	ng	99
16) Anthracene	16.20	178	229084	3.64	ng	100
17) Fluoranthene	18.08	202	262398	3.37	ng	99
19) Pyrene	18.47	202	280413	3.59	ng	100
21) Benzo(a)anthracene	20.19	228	254780	3.24	ng	98
22) Chrysene	20.24	228	292569	3.25	ng	96
23) Indeno(1,2,3-cd)pyrene	23.69	276	257938	3.21	ng	99
25) Benzo(b)fluoranthene	21.54	252	268241	3.23	ng	99
26) Benzo(k)fluoranthene	21.57	252	287715	3.40	ng	99
27) Benzo(a)pyrene	22.00	252	245792	3.21	ng	99
28) Dibenzo(a,h)anthracene	23.64	278	230415	3.08	ng	99
29) Benzo(g,h,i)perylene	24.24	276	262789	3.17	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020516\
Data File : BE091400.D
Acq On : 5 Feb 2016 15:14
Operator : UM/IZ
Sample : PB88200BSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
PB88200BSD

Quant Time: Feb 05 16:45:10 2016
Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Feb 03 00:27:53 2016
Response via : Initial Calibration

