

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020516\
 Data File : BE091402.D
 Acq On : 5 Feb 2016 16:29
 Operator : UM/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 05 17:43:27 2016
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\8270-SIM-BE020216.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Feb 03 00:27:53 2016
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	90	0.00
2 S	2-Fluorophenol	0.896	0.899	-0.3	103	0.00
3 S	Phenol-d6	1.352	1.423	-5.3	100	0.00
4 I	Naphthalene-d8	1.000	1.000	0.0	92	0.00
5 S	Nitrobenzene-d5	0.241	0.253	-5.0	111	0.00
6	Naphthalene	1.128	1.041	7.7	96	0.00
7	2-Methylnaphthalene	0.609	0.597	2.0	102	0.00
8 I	Acenaphthene-d10	1.000	1.000	0.0	93	0.00
9 S	2,4,6-Tribromophenol	0.153	0.147	3.9	114	0.00
10 S	2-Fluorobiphenyl	1.301	1.222	6.1	99	0.00
11	Acenaphthylene	7.133	6.944	2.6	107	0.00
12	Acenaphthene	1.373	1.233	10.2	97	0.00
13	Fluorene	1.772	1.659	6.4	98	0.00
14 I	Phenanthrene-d10	1.000	1.000	0.0	93	0.00
15	Phenanthrene	1.313	1.208	8.0	98	0.00
16	Anthracene	1.100	1.068	2.9	101	0.00
17	Fluoranthene	1.203	1.193	0.8	107	0.00
18 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
19	Pyrene	1.083	1.037	4.2	108	0.00
20 S	Terphenyl-d14	0.740	0.723	2.3	107	0.00
21	Benzo(a)anthracene	0.926	0.886	4.3	113	0.00
22	Chrysene	1.371	1.224	10.7	102	0.00
23	Indeno(1,2,3-cd)pyrene	0.929	0.919	1.1	108	0.00
24 I	Perylene-d12	1.000	1.000	0.0	101	0.00
25	Benzo(b)fluoranthene	1.048	1.027	2.0	108	0.00
26	Benzo(k)fluoranthene	1.187	1.121	5.6	107	0.00
27 C	Benzo(a)pyrene	0.939	0.907	3.4	108	0.00
28	Dibenzo(a,h)anthracene	0.827	0.848	-2.5	114	0.00
29	Benzo(g,h,i)perylene	1.055	1.020	3.3	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0