

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020518\
 Data File : BE095504.D
 Acq On : 5 Feb 2018 23:09
 Operator : SJ/JU
 Sample : J1314-08
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :

Quant Time: Feb 06 02:11:17 2018
 Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 06 02:09:13 2018
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1921	0.40	ng/ul	0.00
2) Naphthalene-d8	11.26	136	5531	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.02	164	3503	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	7975	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	7807	0.40	ng/ul	0.00
20) Perylene-d12	24.26	264	342390	0.40	ng/ul	-0.16
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	2802	0.31	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	7926	0.26	ng/ul	0.00
Target Compounds				Ovalue		
11) Pentachlorophenol	17.36	266	401	0.260	ng/ul	98
12) Phenanthrene	17.75	178	890	0.033	ng/ul#	59
17) Pyrene	20.05	202	1718	0.066	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA E\DATA\BE020518\
Data File : BE095504.D
Acq On : 5 Feb 2018 23:09
Operator : SJ/JU
Sample : J1314-08
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :

Quant Time: Feb 06 02:11:17 2018
Quant Method : Z:\HPCHEM1\BNA E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 02:09:13 2018
Response via : Initial Calibration









#11
 Pentachlorophenol
 Concen: 0.260 ng/ul
 RT: 17.36 min Scan# 1997
 Delta R.T. 0.00 min
 Lab File: BE095504.D
 Acq: 5 Feb 2018 23:09




Instrument :
 BNA_E
 ClientSampleId :




Tot Ion:266 Resp: 401
 Ion Ratio Lower Upper
 266 100
 264 63.1 47.9 71.9
 268 62.8 49.8 74.8



#12
 Phenanthrene
 Concen: 0.033 ng/ul
 RT: 17.75 min Scan# 2014
 Delta R.T. 0.00 min
 Lab File: BE095504.D
 Acq: 5 Feb 2018 23:09

Tgt Ion:178 Resp: 890
 Ion Ratio Lower Upper
 178 100
 179 43.8 18.4 27.6#
 176 34.2 13.6 20.4#





#17

Pvrene

Concen: 0.066 ng/ul

RT: 20.05 min Scan# 2460

Delta R.T. -0.02 min

Lab File: BE095504.D

Acq: 5 Feb 2018 23:09

Instrument :

BNA_E

ClientSampleId :

Tot Ion: 202 Resp: 1718

Ion Ratio Lower Upper

202 100

101 141.5 18.2 27.4#

100 859.7 15.6 23.4#

