

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020518\
 Data File : BE095525.D
 Acq On : 6 Feb 2018 11:38
 Operator : SJ/JU
 Sample : J1296-02DL 5X
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 F6C20DL

Quant Time: Feb 06 12:30:02 2018
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 06 04:22:42 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	8.43	152	1719	0.40	ng/ul	0.00
2) Naphthalene-d8	11.25	136	4792	0.40	ng/ul	0.00
6) Acenaphthene-d10	15.01	164	2901	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.71	188	7187	0.40	ng/ul	0.00
16) Chrysene-d12	21.80	240	7007	0.40	ng/ul	0.00
20) Perylene-d12	24.43	264	7124	0.40	ng/ul	0.00
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	12.81	152	666	0.09	ng/ul	0.00
14) Fluoranthene-d10	19.69	212	1734	0.06	ng/ul	0.00
Target Compounds						
3) Naphthalene	11.30	128	1244	0.093	ng/ul#	83
11) Pentachlorophenol	17.36	266	1511	1.085	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : \\74.0.250.170\SV0ASRV\HPCHEM1\BNA_E\DATA\BE020518\
Data File : BE095525.D
Acq On : 6 Feb 2018 11:38
Operator : SJ/JU
Sample : J1296-02DL 5X
Misc :
ALS Vial : 41 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
F6C20DL

Quant Time: Feb 06 12:30:02 2018
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SOM-EPA-SIM-BE012618.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 06 04:22:42 2018
Response via : Initial Calibration

