

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089527.D
 Acq On : 6 Feb 2015 14:31
 Operator : TP/IZ
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICC0.8

Quant Time: Feb 07 03:34:39 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 03:25:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.43	152	38016	5.00	ng	0.00
2) Naphthalene-d8	12.33	136	145367	5.00	ng	0.00
6) Acenaphthene-d10	16.51	164	59245	5.00	ng	0.00
11) Phenanthrene-d10	19.85	188	108442	5.00	ng	0.00
16) Chrysene-d12	23.72	240	126518	5.00	ng	0.00
22) Perylene-d12	25.42	264	137832	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.69	82	7143	0.69	ng	0.00
7) 2-Fluorobiphenyl	14.97	172	11983	0.64	ng	0.00
18) Terphenyl-d14	22.36	244	12928	0.57	ng	0.00

Target Compounds					Qvalue	
4) Naphthalene	12.38	128	25414	0.71	ng	99
5) 2-Methylnaphthalene	14.04	142	14287	0.63	ng	100
8) Acenaphthylene	16.16	152	75989	0.61	ng	100
9) Acenaphthene	16.59	154	12104	0.63	ng	99
10) Fluorene	17.84	166	12929	0.60	ng	99
12) Pentachlorophenol	19.52	266	256	0.31	ng	95
13) Phenanthrene	19.89	178	20577	0.63	ng	99
14) Anthracene	19.99	178	17156	0.58	ng	100
15) Fluoranthene	21.79	202	19276	0.64	ng	100
17) Pyrene	22.11	202	20504	0.53	ng	99
19) Benzo(a)anthracene	23.71	228	85436	0.62	ng	97
20) Chrysene	23.75	228	88075	0.62	ng	99
21) Indeno(1,2,3-cd)pyrene	26.85	276	114606	0.72	ng	99
23) Benzo(b)fluoranthene	24.98	252	20580	0.67	ng	99
24) Benzo(k)fluoranthene	25.01	252	25843	0.59	ng	99
25) Benzo(a)pyrene	25.35	252	19867	0.62	ng	98
26) Dibenzo(a,h)anthracene	26.89	278	24291	0.68	ng	99
27) Benzo(g,h,i)perylene	27.28	276	97960	0.66	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
Data File : BE089527.D
Acq On : 6 Feb 2015 14:31
Operator : TP/IZ
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICC0.8

Quant Time: Feb 07 03:34:39 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 07 03:25:41 2015
Response via : Initial Calibration

