

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089528.D
 Acq On : 6 Feb 2015 15:10
 Operator : TP/IZ
 Sample : SSTDICCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDICCC001

Quant Time: Feb 07 03:35:02 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 03:25:41 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.42	152	33996	5.00	ng	0.00
2) Naphthalene-d8	12.33	136	128329	5.00	ng	0.00
6) Acenaphthene-d10	16.51	164	48484	5.00	ng	0.00
11) Phenanthrene-d10	19.85	188	80109	5.00	ng	0.00
16) Chrysene-d12	23.72	240	91063	5.00	ng	0.00
22) Perylene-d12	25.42	264	107033	5.00	ng	0.00

System Monitoring Compounds						
3) Nitrobenzene-d5	10.69	82	7980	0.88	ng	0.00
7) 2-Fluorobiphenyl	14.96	172	12667	0.83	ng	0.00
18) Terphenyl-d14	22.36	244	11528	0.71	ng	0.00

Target Compounds					Qvalue	
4) Naphthalene	12.38	128	28080	0.89	ng	100
5) 2-Methylnaphthalene	14.04	142	15569	0.78	ng	100
8) Acenaphthylene	16.16	152	79300	0.78	ng	100
9) Acenaphthene	16.59	154	12627	0.81	ng	100
10) Fluorene	17.84	166	12707	0.72	ng	100
12) Pentachlorophenol	19.52	266	206	0.34	ng	100
13) Phenanthrene	19.89	178	19276	0.80	ng	100
14) Anthracene	19.99	178	16018	0.73	ng	100
15) Fluoranthene	21.79	202	17442	0.78	ng	100
17) Pyrene	22.11	202	18427	0.67	ng	100
19) Benzo(a)anthracene	23.71	228	76051	0.77	ng	100
20) Chrysene	23.76	228	81383	0.79	ng	100
21) Indeno(1,2,3-cd)pyrene	26.85	276	120393	1.06	ng	100
23) Benzo(b)fluoranthene	24.98	252	19189	0.81	ng	100
24) Benzo(k)fluoranthene	25.01	252	24775	0.73	ng	100
25) Benzo(a)pyrene	25.35	252	19608	0.79	ng	100
26) Dibenzo(a,h)anthracene	26.89	278	25873	0.94	ng	100
27) Benzo(g,h,i)perylene	27.28	276	105350	0.91	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
Data File : BE089528.D
Acq On : 6 Feb 2015 15:10
Operator : TP/IZ
Sample : SSTDICCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDICCC001

Quant Time: Feb 07 03:35:02 2015
Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 07 03:25:41 2015
Response via : Initial Calibration

