

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089539.D
 Acq On : 6 Feb 2015 23:01
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 07 05:07:25 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.42	152	32758	5.00	ng	0.00
2) Naphthalene-d8	12.33	136	119736	5.00	ng	0.00
6) Acenaphthene-d10	16.51	164	42066	5.00	ng	0.00
11) Phenanthrene-d10	19.84	188	62815	5.00	ng	0.00
16) Chrysene-d12	23.72	240	67492	5.00	ng	0.00
22) Perylene-d12	25.41	264	82426	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) Nitrobenzene-d5	10.69	82	7522	0.95	ng	0.00
7) 2-Fluorobiphenyl	14.96	172	11345	1.01	ng	0.00
18) Terphenyl-d14	22.36	244	7848	0.88	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Naphthalene	12.38	128	26038	0.96	ng	99
5) 2-Methylnaphthalene	14.03	142	14259	0.93	ng	97
8) Acenaphthylene	16.16	152	68636	0.97	ng	100
9) Acenaphthene	16.59	154	10688	0.95	ng	97
10) Fluorene	17.84	166	10447	0.88	ng	100
12) Pentachlorophenol	19.52	266	50	0.64	ng	93
13) Phenanthrene	19.89	178	14873	0.96	ng	99
14) Anthracene	19.98	178	12356	0.94	ng	99
15) Fluoranthene	21.79	202	12357	0.85	ng	99
17) Pyrene	22.10	202	13017	0.91	ng	100
19) Benzo(a)anthracene	23.70	228	51160	0.85	ng	92
20) Chrysene	23.75	228	61075	1.00	ng	93
21) Indeno(1,2,3-cd)pyrene	26.85	276	89359	1.05	ng	99
23) Benzo(b)fluoranthene	24.98	252	12405	0.78	ng	97
24) Benzo(k)fluoranthene	25.00	252	19407	0.98	ng	98
25) Benzo(a)pyrene	25.35	252	13392	0.84	ng	99
26) Dibenzo(a,h)anthracene	26.88	278	18249	0.93	ng	100
27) Benzo(g,h,i)perylene	27.28	276	80803	1.01	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089539.D
 Acq On : 6 Feb 2015 23:01
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
LabSampleId :
 SSTDCCC001

Quant Time: Feb 07 05:07:25 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

