

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089541.D
 Acq On : 7 Feb 2015 00:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 07 05:55:29 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.42	152	30887	5.00	ng	0.00
2) Naphthalene-d8	12.33	136	113601	5.00	ng	0.00
6) Acenaphthene-d10	16.51	164	41746	5.00	ng	0.00
11) Phenanthrene-d10	19.85	188	69561	5.00	ng	0.00
16) Chrysene-d12	23.72	240	85278	5.00	ng	0.00
22) Perylene-d12	25.41	264	98282	5.00	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
3) Nitrobenzene-d5	10.69	82	6796	0.90	ng	0.00
7) 2-Fluorobiphenyl	14.96	172	11103	1.00	ng	0.00
18) Terphenyl-d14	22.36	244	9852	0.88	ng	0.00

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Naphthalene	12.38	128	24729	0.97	ng	99
5) 2-Methylnaphthalene	14.03	142	13586	0.93	ng	98
8) Acenaphthylene	16.16	152	67036	0.96	ng	100
9) Acenaphthene	16.59	154	10905	0.98	ng	100
10) Fluorene	17.84	166	10838	0.92	ng	99
12) Pentachlorophenol	19.52	266	88	0.70	ng	93
13) Phenanthrene	19.89	178	16611	0.96	ng	98
14) Anthracene	19.98	178	13628	0.94	ng	99
15) Fluoranthene	21.79	202	14824	0.92	ng	99
17) Pyrene	22.10	202	16244	0.89	ng	99
19) Benzo(a)anthracene	23.70	228	65226	0.86	ng	95
20) Chrysene	23.75	228	76994	1.00	ng	94
21) Indeno(1,2,3-cd)pyrene	26.85	276	96809	0.90	ng	97
23) Benzo(b)fluoranthene	24.97	252	15074	0.79	ng	99
24) Benzo(k)fluoranthene	25.00	252	22423	0.95	ng	99
25) Benzo(a)pyrene	25.34	252	15480	0.82	ng	99
26) Dibenzo(a,h)anthracene	26.88	278	20634	0.89	ng	97
27) Benzo(g,h,i)perylene	27.28	276	87786	0.92	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
Data File : BE089541.D
Acq On : 7 Feb 2015 00:58
Operator : TP/IZ
Sample : SSTDCCC001
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_E
LabSampleId :
SSTDCCC001

Quant Time: Feb 07 05:55:29 2015
Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 07 04:07:57 2015
Response via : Initial Calibration

