

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
 Data File : BE089544.D
 Acq On : 7 Feb 2015 3:33
 Operator : TP/IZ
 Sample : G1238-02
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 MW2

Manual Integrations
 APPROVED

apatel
 2/9/2015 11:54:58 AM

Quant Time: Feb 07 06:08:48 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	9.42	152	36337	5.00	ng	0.00
2) Naphthalene-d8	12.33	136	141032	5.00	ng	0.00
6) Acenaphthene-d10	16.51	164	60022	5.00	ng	0.00
11) Phenanthrene-d10	19.84	188	114576	5.00	ng	0.00
16) Chrysene-d12	23.72	240	130664	5.00	ng	0.00
22) Perylene-d12	25.41	264	138344	5.00	ng	0.00
System Monitoring Compounds						
3) Nitrobenzene-d5	10.69	82	64933	6.93	ng	0.00
7) 2-Fluorobiphenyl	14.96	172	114665	7.16	ng	0.00
18) Terphenyl-d14	22.36	244	162729	9.44	ng	0.00
Target Compounds						Qvalue
13) Phenanthrene	19.89	178	263	0.01	ng	95
14) Anthracene	19.98	178	164m	0.01	ng	
15) Fluoranthene	21.79	202	320	0.01	ng	# 92
17) Pyrene	22.10	202	370	0.01	ng	# 92
19) Benzo(a)anthracene	23.71	228	3607	0.03	ng	# 87
20) Chrysene	23.75	228	2849	0.02	ng	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\HPCHEM1\BNA_E\DATA\BE020615\
Data File : BE089544.D
Acq On : 7 Feb 2015 3:33
Operator : TP/IZ
Sample : G1238-02
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_E
ClientSampleId :
MW2

Manual Integrations
APPROVED

apatel
2/9/2015 11:54:58 AM

Quant Time: Feb 07 06:08:48 2015
Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Feb 07 04:07:57 2015
Response via : Initial Calibration

