

Data Path : Z:\HPCHEM1\BNA_E\Data\BE020615\
 Data File : BE089532.D
 Acq On : 6 Feb 2015 18:27
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE020615

Quant Time: Feb 09 10:50:34 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	99	0.00
2 I	Naphthalene-d8	5.000	5.000	0.0	98	0.00
3 S	Nitrobenzene-d5	1.000	0.923	7.7	96	0.00
4	Naphthalene	1.000	0.971	2.9	98	0.00
5	2-Methylnaphthalene	1.000	0.946	5.4	97	0.00
6 I	Acenaphthene-d10	5.000	5.000	0.0	99	0.00
7 S	2-Fluorobiphenyl	1.000	0.978	2.2	99	0.00
8	Acenaphthylene	1.000	0.981	1.9	99	0.00
9	Acenaphthene	1.000	0.968	3.2	98	0.00
10	Fluorene	1.000	0.943	5.7	101	0.00
11 I	Phenanthrene-d10	5.000	5.000	0.0	105	0.00
12	Pentachlorophenol	1.000	0.788	21.2	79	0.00
13	Phenanthrene	1.000	0.975	2.5	105	0.00
14	Anthracene	1.000	0.952	4.8	104	0.00
15	Fluoranthene	1.000	0.963	3.7	107	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	109	0.00
17	Pyrene	1.000	0.936	6.4	108	0.00
18 S	Terphenyl-d14	1.000	0.965	3.5	110	0.00
19	Benzo(a)anthracene	1.000	0.924	7.6	108	0.00
20	Chrysene	1.000	0.989	1.1	109	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	0.925	7.5	96	0.00
22 I	Perylene-d12	5.000	5.000	0.0	103	0.00
23	Benzo(b)fluoranthene	1.000	0.892	10.8	100	0.00
24	Benzo(k)fluoranthene	1.000	0.974	2.6	105	0.00
25 C	Benzo(a)pyrene	1.000	0.906	9.4	99	0.00
26	Dibenzo(a,h)anthracene	1.000	0.939	6.1	95	0.00
27	Benzo(g,h,i)perylene	1.000	0.947	5.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0