

Data Path : Z:\HPCHEM1\BNA_E\Data\BE020615\
 Data File : BE089532.D
 Acq On : 6 Feb 2015 18:27
 Operator : TP/IZ
 Sample : SSTDICV001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 ClientSampleId :
 ICSVBE020615

Quant Time: Feb 09 10:50:34 2015
 Quant Method : Z:\HPCHEM1\BNA_E\METHODS\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
3 S	Nitrobenzene-d5	0.332	0.307	7.5	96	0.00
4	Naphthalene	1.127	1.094	2.9	98	0.00
5	2-Methylnaphthalene	0.640	0.605	5.5	97	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
7 S	2-Fluorobiphenyl	1.334	1.305	2.2	99	0.00
8	Acenaphthylene	8.406	8.243	1.9	99	0.00
9	Acenaphthene	1.335	1.293	3.1	98	0.00
10	Fluorene	1.417	1.336	5.7	101	0.00
11 I	Phenanthrene-d10	1.000	1.000	0.0	105	0.00
12	Pentachlorophenol	0.030	0.010	66.7#	79	0.00
13	Phenanthrene	1.239	1.208	2.5	105	0.00
14	Anthracene	1.046	0.997	4.7	104	0.00
15	Fluoranthene	1.154	1.112	3.6	107	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	109	0.00
17	Pyrene	1.065	0.997	6.4	108	0.00
18 S	Terphenyl-d14	0.660	0.637	3.5	110	0.00
19	Benzo(a)anthracene	4.449	4.113	7.6	108	0.00
20	Chrysene	4.519	4.469	1.1	109	0.00
21	Indeno(1,2,3-cd)pyrene	6.282	5.814	7.4	96	0.00
22 I	Perylene-d12	1.000	1.000	0.0	103	0.00
23	Benzo(b)fluoranthene	0.971	0.866	10.8	100	0.00
24	Benzo(k)fluoranthene	1.207	1.176	2.6	105	0.00
25 C	Benzo(a)pyrene	0.966	0.875	9.4	99	0.00
26	Dibenzo(a,h)anthracene	1.185	1.113	6.1	95	0.00
27	Benzo(g,h,i)perylene	4.838	4.583	5.3	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0