

Data Path : Z:\HPCHEM1\BNA_E\Data\BE020615\
 Data File : BE089541.D
 Acq On : 7 Feb 2015 00:58
 Operator : TP/IZ
 Sample : SSTDCCC001
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_E
 LabSampleId :
 SSTDCCC001

Quant Time: Feb 07 05:55:29 2015
 Quant Method : Z:\HPCHEM1\BNA_E\methods\SIMPAH-BE020615.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Feb 07 04:07:57 2015
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	91	0.00
2 I	Naphthalene-d8	5.000	5.000	0.0	89	0.00
3 S	Nitrobenzene-d5	1.000	0.901	9.9	85	0.00
4	Naphthalene	1.000	0.966	3.4	88	0.00
5	2-Methylnaphthalene	1.000	0.935	6.5	87	0.00
6 I	Acenaphthene-d10	5.000	5.000	0.0	86	0.00
7 S	2-Fluorobiphenyl	1.000	0.997	0.3	88	0.00
8	Acenaphthylene	1.000	0.955	4.5	85	0.00
9	Acenaphthene	1.000	0.978	2.2	86	0.00
10	Fluorene	1.000	0.916	8.4	85	0.00
11 I	Phenanthrene-d10	5.000	5.000	0.0	87	0.00
12	Pentachlorophenol	1.000	0.699	30.1#	43	0.00
13	Phenanthrene	1.000	0.964	3.6	86	0.00
14	Anthracene	1.000	0.936	6.4	85	0.00
15	Fluoranthene	1.000	0.923	7.7	85	0.00
16 I	Chrysene-d12	5.000	5.000	0.0	94	0.00
17	Pyrene	1.000	0.894	10.6	88	0.00
18 S	Terphenyl-d14	1.000	0.876	12.4	85	0.00
19	Benzo(a)anthracene	1.000	0.860	14.0	86	0.00
20	Chrysene	1.000	0.999	0.1	95	0.00
21	Indeno(1,2,3-cd)pyrene	1.000	0.903	9.7	80	0.00
22 I	Perylene-d12	5.000	5.000	0.0	92	0.00
23	Benzo(b)fluoranthene	1.000	0.790	21.0	79	0.00
24	Benzo(k)fluoranthene	1.000	0.945	5.5	91	0.00
25 C	Benzo(a)pyrene	1.000	0.816	18.4	79	0.00
26	Dibenzo(a,h)anthracene	1.000	0.886	11.4	80	0.00
27	Benzo(g,h,i)perylene	1.000	0.923	7.7	83	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0